



Certificate:
0000025926_03 / 19 August 2011

Calculation of overall uncertainty for QAL1 in EN 14181 and EN 15267-3

Manufacturer data

Manufacturer: Sick Maihak GmbH
 Name of measuring system: MCS 100 FT
 Serial Number: TUEV 1, TUEV 2, TUEV 3, TUEV 4
 Measuring Principle: FTIR

TÜV Data

Approval Report: 936/21206925A / 2008-10-20

Editor: C. Landgraf

Date: 2009-10-26

Measurement Component

HCl: 15 mg/m³

Evaluation of the cross sensitivity (CS)

Sum of positive CS at zero point	0.59 mg/m³
Sum of negative CS at zero point	0.08 mg/m³
Sum of positive CS at reference point	0.50 mg/m³
Sum of negative CS at reference point	0.08 mg/m³
Maximum sum of cross sensitivities	0.59 mg/m³
Uncertainty of cross sensitivity	0.34 mg/m³

Calculation of the combined standard uncertainty

Test Value

Standard deviation from paired measurements under field conditions *

	<i>u</i>	<i>u</i> ²
<i>u_D</i>	0.170 mg/m³	0.029 (mg/m³) ²
<i>u_{lof}</i>	0.170 mg/m³	0.029 (mg/m³) ²
<i>u_{dz}</i>	-0.210 mg/m³	0.044 (mg/m³) ²
<i>u_{ds}</i>	-0.250 mg/m³	0.063 (mg/m³) ²
<i>u_t</i>	-0.300 mg/m³	0.090 (mg/m³) ²
<i>u_v</i>	0.060 mg/m³	0.004 (mg/m³) ²
<i>u_i</i>	0.341 mg/m³	0.116 (mg/m³) ²
<i>u_p</i>	0.000 mg/m³	0.000 (mg/m³) ²
<i>u_m</i>	0.121 mg/m³	0.015 (mg/m³) ²

* The bigger value of "Repeatability standard deviation at span" or
 "Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (*u_c*)

$$u_c = \sqrt{\sum (u_{max,i})^2}$$

$$U = u_c * k = u_c * 1,96$$

0.62 mg/m³
 1.22 mg/m³

Total expanded uncertainty

Relative total expanded uncertainty

<i>U</i> in % of the ELV 10 mg/m³	12.2
<i>U</i> in % of the ELV 10 mg/m³	40.0
<i>U</i> in % of the ELV 10 mg/m³	30.0

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3



Calculation of overall uncertainty for QAL1 in EN 14181 and EN 15267-3

Manufacturer data

Manufacturer

Name of measuring system

Serial Number

Measuring Principle

Sick Maihak GmbH

MCS 100 FT

TUEV 1, TUEV 2, TUEV 3, TUEV 4

FTIR

TÜV Data

Approval Report

936/21206925A / 2008-10-20

Editor

C. Landgraf

Date

2009-10-26

Measurement Component

Certificated range

HF

3 mg/m³

Evaluation of the cross sensitivity (CS)

Sum of positive CS at zero point

0.12 mg/m³

Sum of negative CS at zero point

-0.08 mg/m³

Sum of positive CS at reference point

0.05 mg/m³

Sum of negative CS at reference point

-0.11 mg/m³

Maximum sum of cross sensitivities

0.12 mg/m³

Uncertainty of cross sensitivity

0.07 mg/m³

Calculation of the combined standard uncertainty

Test Value

Repeatability standard deviation at set point *

	u	u²
u _r	0.050 mg/m³	0.003 (mg/m³)²
u _{lof}	-0.029 mg/m³	0.001 (mg/m³)²
u _{d,z}	-0.059 mg/m³	0.003 (mg/m³)²
u _{d,s}	-0.068 mg/m³	0.005 (mg/m³)²
u _t	0.081 mg/m³	0.007 (mg/m³)²
u _v	0.023 mg/m³	0.001 (mg/m³)²
u _j	0.069 mg/m³	0.005 (mg/m³)²
u _p	0.000 mg/m³	0.000 (mg/m³)²
u _m	0.024 mg/m³	0.001 (mg/m³)²

Lack of fit

Zero drift from field test

Span drift from field test

Influence of ambient temperature at span

Influence of supply voltage

Cross sensitivity (interference)

Influence of sample gas flow

Uncertainty of reference material at 70% of certification range

* The bigger value of: "Repeatability standard deviation at span" or "Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,i})^2}$$

0.15 mg/m³

Total expanded uncertainty

0.30 mg/m³

Relative total expanded uncertainty

U in % of the ELV 1 mg/m³ 30.3

Requirement of 2000/76/EC and 2001/80/EC

U in % of the ELV 1 mg/m³ 40.0

Requirement of EN 15267-3

U in % of the ELV 1 mg/m³ 30.0



Calculation of overall uncertainty for QAL1 in EN 14181 and EN 15267-3

Manufacturer data

Manufacturer: SICK Maihak GmbH.
 Name of measuring system: MCS 100 FT
 Serial Number: TUEV 1, TUEV 2, TUEV 3, TUEV 4
 Measuring Principle: FTIR

TÜV Data

Approval Report: 936/21206925A / 2008-10-20
 Editor: C. Landgraf
 Date: 2009-10-26

Measurement Component

Certified range: CH₄ 50 mg/m³

Evaluation of the cross sensitivity (CS)

Sum of positive CS at zero point	0.55 mg/m ³
Sum of negative CS at zero point	0.25 mg/m ³
Sum of positive CS at reference point	1.35 mg/m ³
Sum of negative CS at reference point	-0.60 mg/m ³
Maximum sum of cross sensitivities	1.35 mg/m ³
Uncertainty of cross sensitivity	0.78 mg/m ³

Calculation of the combined standard uncertainty

Test Value

Test Value	u	u ²
Standard deviation from paired measurements under field conditions *	0.540 mg/m ³	0.292 (mg/m ³) ²
Lack of fit	-0.200 mg/m ³	0.040 (mg/m ³) ²
Zero drift from field test	-0.720 mg/m ³	0.518 (mg/m ³) ²
Span drift from field test	-0.870 mg/m ³	0.757 (mg/m ³) ²
Influence of ambient temperature at span	0.400 mg/m ³	0.160 (mg/m ³) ²
Influence of supply voltage	0.060 mg/m ³	0.004 (mg/m ³) ²
Cross sensitivity (interference)	0.779 mg/m ³	0.608 (mg/m ³) ²
Influence of sample gas flow	0.000 mg/m ³	0.000 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	0.404 mg/m ³	0.163 (mg/m ³) ²

* The bigger value of: "Repeatability standard deviation at span" or
 "Standard deviation from paired measurements under field conditions"

$$\text{Combined standard uncertainty (u_c)} = \sqrt{\sum (u_{\text{max}, i})^2} \quad 1.59 \text{ mg/m}^3$$

$$\text{Total expanded uncertainty} = u_c * k = u_c * 1.96 \quad 3.12 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC**	U in % of the ELV 20 mg/m ³	15.6
Requirement of EN 15267-3	U in % of the ELV 20 mg/m ³	30.0

U in % of the ELV 20 mg/m³

15.6

U in % of the ELV 20 mg/m³

30.0

U in % of the ELV 20 mg/m³

22.5

** For this component no requirements in the EC-directives 2001/80/EC und 2000/76/EC are given.

The chosen value is recommended by the certification body.



Calculation of overall uncertainty for QAL1 in EN 14181 and EN 15267-3

Manufacturer data

Manufacturer Sick Maihak GmbH
 Name of measuring system MCS 100 FT
 Serial Number TUEV 1, TUEV 2, TUEV 3, TUEV 4
 Measuring Principle FTIR

TÜV Data

Approval Report 936/21206925A / 2008-10-20
 Editor C. Landgraf
 Date 2009-10-26

Measurement Component

Certified range CO₂ 25 Vol.-%

Evaluation of the cross sensitivity (CS)

Sum of positive CS at zero point	0.23 Vol.-%
Sum of negative CS at zero point	-0.73 Vol.-%
Sum of positive CS at reference point	0.80 Vol.-%
Sum of negative CS at reference point	-0.78 Vol.-%
Maximum sum of cross sensitivities	0.80 Vol.-%
Uncertainty of cross sensitivity	0.46 Vol.-%

Calculation of the combined standard uncertainty

Test Value	u	u²
Standard deviation from paired measurements under field conditions *	u _D 0.360 Vol.-%	0.130 (Vol.-%) ²
Lack of fit	u _{lof} 0.100 Vol.-%	0.010 (Vol.-%) ²
Zero drift from field test	u _{d,z} 0.300 Vol.-%	0.090 (Vol.-%) ²
Span drift from field test	u _{d,s} 0.390 Vol.-%	0.152 (Vol.-%) ²
Influence of ambient temperature at span	u _t 0.300 Vol.-%	0.090 (Vol.-%) ²
Influence of supply voltage	u _v 0.060 Vol.-%	0.004 (Vol.-%) ²
Cross sensitivity (interference)	u _i 0.462 Vol.-%	0.213 (Vol.-%) ²
Influence of sample gas flow	u _p 0.000 Vol.-%	0.000 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u _m 0.202 Vol.-%	0.041 (Vol.-%) ²

* The bigger value of: "Repeatability standard deviation at span" or
 "Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,i})^2}$	0.85 Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1,96$	1.67 Vol.-%

Relative total expanded uncertainty

U in % of the range 25 Vol.-% 6.7

Requirement of 2000/76/EC and 2001/80/EC**

U in % of the range 25 Vol.-% 10.0

Requirement of EN 15267-3

U in % of the range 25 Vol.-% 7.5

** For this component no requirements in the EC-directives 2001/80/EC und 2000/76/EC are given.

The chosen value is recommended by the certification body.



Calculation of overall uncertainty for QAL1 in EN 14181 and EN 15267-3.

Manufacturer data

Manufacturer: SICK Maljak GmbH
 Name of measuring system: MCS 100 FT
 Serial Number: TUEV 1, TUEV 2, TUEV 3, TUEV 4
 Measuring Principle: FTIR

TÜV Data

Approval Report: 936/21206925A / 2008-10-20
 Editor: C. Landgraf
 Date: 2009-10-26

Measurement Component

Certified range: H₂O 40 Vol.-%

Evaluation of the cross sensitivity (CS)

Sum of positive CS at zero point	0.80 Vol.-%
Sum of negative CS at zero point	-0.20 Vol.-%
Sum of positive CS at reference point	0.76 Vol.-%
Sum of negative CS at reference point	-0.76 Vol.-%
Maximum sum of cross sensitivities	0.80 Vol.-%
Uncertainty of cross sensitivity	0.46 Vol.-%

Calculation of the combined standard uncertainty

Test Value	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.160 Vol.-%	0.026 (Vol.-%) ²
Lack of fit	u _{lf} 0.370 Vol.-%	0.137 (Vol.-%) ²
Zero drift from field test	u _{d,z} -0.600 Vol.-%	0.360 (Vol.-%) ²
Span drift from field test	u _{d,s} 0.670 Vol.-%	0.449 (Vol.-%) ²
Influence of ambient temperature at span	u _t 0.280 Vol.-%	0.078 (Vol.-%) ²
Influence of supply voltage	u _v 0.050 Vol.-%	0.003 (Vol.-%) ²
Cross sensitivity (interference)	u _i 0.462 Vol.-%	0.213 (Vol.-%) ²
Influence of sample gas flow	u _p 0.000 Vol.-%	0.000 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u _m 0.323 Vol.-%	0.105 (Vol.-%) ²

* The bigger value of: "Repeatability standard deviation at span" or
 "Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,i})^2}$$

1.17 Vol.-%

2.29 Vol.-%

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96$$

Relative total expanded uncertainty
 Requirement of 2000/76/EC and 2001/80/EC**
 Requirement of EN 15267-3

U in % of the range 40 Vol.-%	5.7
U in % of the range 40 Vol.-%	10.0
U in % of the range 40 Vol.-%	7.5

** For this component no requirements in the EC-directives 2001/80/EC und 2000/76/EC are given.

The chosen value is recommended by the certification body.



Calculation of overall uncertainty for QAL1 in EN 14181 and EN 15267-3

Manufacturer data

Manufacturer

Name of measuring system

Serial Number

Measuring Principle

Sick Maihak GmbH.

MCS 100 FT

TUEV 1, TUEV 2, TUEV 3, TUEV 4

FTIR

TÜV Data

Approval Report

936/21206925A / 2008-10-20

Editor

C. Landgraf

Date

2009-10-26

Measurement Component

Certificated range

N₂O

50 mg/m³

Evaluation of the cross sensitivity (CS)

Sum of positive CS at zero point

1.95 mg/m³

Sum of negative CS at zero point

-0.70 mg/m³

Sum of positive CS at reference point

1.75 mg/m³

Sum of negative CS at reference point

-0.80 mg/m³

Maximum sum of cross sensitivities

1.95 mg/m³

Uncertainty of cross sensitivity

1.13 mg/m³

Calculation of the combined standard uncertainty

Test Value

Repeatability standard deviation at set point *

	u	u ²
u _r	0.250 mg/m ³	0.063 (mg/m ³) ²
u _{lof}	0.140 mg/m ³	0.020 (mg/m ³) ²
u _{d,z}	-0.120 mg/m ³	0.014 (mg/m ³) ²
u _{d,s}	-0.520 mg/m ³	0.270 (mg/m ³) ²
u _t	-0.320 mg/m ³	0.102 (mg/m ³) ²
u _v	0.120 mg/m ³	0.014 (mg/m ³) ²
u _i	1.126 mg/m ³	1.268 (mg/m ³) ²
u _p	0.000 mg/m ³	0.000 (mg/m ³) ²
u _m	0.404 mg/m ³	0.163 (mg/m ³) ²

* The bigger value of: "Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,i})^2} \quad 1.38 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 2.71 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 20 mg/m³ 13.6

Requirement of 2000/76/EC and 2001/80/EC**

20.0

Requirement of EN 15267-3

U in % of the ELV 20 mg/m³ 15.0

** For this component no requirements in the EC-directives 2001/80/EC und 2000/76/EC are given.

The chosen value is recommended by the certification body.



Calculation of overall uncertainty for QAL1 in EN 14181 and EN 15267-3

Manufacturer data

Manufacturer SICK MAIHAK GmbH
 Name of measuring system MCS 100 FT
 Serial Number TUEV 1, TUEV 2, TUEV 3, TUEV 4
 Measuring Principle FTIR

TÜV Data

Approval Report 936/21210511/A / 2010-03-22

Editor Steinhagen
 Date 2010-03-01

Measurement Component

Certified range NH₃ 10 mg/m³

Evaluation of the cross sensitivity (CS)

Sum of positive CS at zero point	0.40 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	0.00 mg/m ³
Sum of negative CS at reference point	-0.29 mg/m ³
Maximum sum of cross sensitivities	0.40 mg/m ³
Uncertainty of cross sensitivity	0.23 mg/m ³

Calculation of the combined standard uncertainty

Test Value

Test Value	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.076 mg/m ³	0.006 (mg/m ³) ²
Lack of fit	u _{for} -0.035 mg/m ³	0.001 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.030 mg/m ³	0.001 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.170 mg/m ³	0.029 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.072 mg/m ³	0.005 (mg/m ³) ²
Influence of supply voltage	u _v 0.072 mg/m ³	0.005 (mg/m ³) ²
Cross sensitivity (interference)	u _i 0.231 mg/m ³	0.053 (mg/m ³) ²
Influence of sample gas flow	u _p 0.000 mg/m ³	0.000 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.081 mg/m ³	0.007 (mg/m ³) ²

* The bigger value of: "Repeatability standard deviation at span" or "Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max})^2}$$

$$u_c = u_{c} * k = u_c * 1.96$$

0.33 mg/m³

0.64 mg/m³

Total expanded uncertainty

Relative total expanded uncertainty

U in % of the range mg/m³ 6.4

Requirement of 2000/76/EC and 2001/80/EC**

U in % of the range mg/m³ 40.0

Requirement of EN 15267-3

U in % of the range mg/m³ 30.0

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

The chosen value is recommended by the certification body.



Calculation of overall uncertainty for QAL1 in EN 14181 and EN 15267-3

Manufacturer data

Manufacturer SICK MAIHAK GmbH
 Name of measuring system MCS 100 FT
 Serial Number TUEV 3, TUEV 4
 Measuring Principle FID

TÜV Data

Approval Report 936/21210511/A / 2010-03-22
 Editor Steinhagen
 Date 2010-03-01

Measurement Component

Certificated range TOC 15 mg/m³

Evaluation of the cross sensitivity (CS)

Sum of positive CS at zero point	0.46 mg/m³
Sum of negative CS at zero point	0.00 mg/m³
Sum of positive CS at reference point	0.26 mg/m³
Sum of negative CS at reference point	0.00 mg/m³
Maximum sum of cross sensitivities	0.46 mg/m³
Uncertainty of cross sensitivity	0.27 mg/m³

Calculation of the combined standard uncertainty

Test Value

Standard deviation from paired measurements under field conditions *

	<i>u</i>	<i>u</i> ²
<i>u_D</i>	0.046 mg/m³	0.002 (mg/m³)²
<i>u_{lof}</i>	0.058 mg/m³	0.003 (mg/m³)²
<i>u_{d,z}</i>	0.152 mg/m³	0.023 (mg/m³)²
<i>u_{d,s}</i>	0.244 mg/m³	0.060 (mg/m³)²
<i>u_t</i>	0.100 mg/m³	0.010 (mg/m³)²
<i>u_v</i>	0.053 mg/m³	0.003 (mg/m³)²
<i>u_i</i>	0.270 mg/m³	0.073 (mg/m³)²
<i>u_p</i>	-0.063 mg/m³	0.004 (mg/m³)²
<i>u_{rm}</i>	0.121 mg/m³	0.015 (mg/m³)²
<i>u_{rf}</i>	0.980 mg/m³	0.960 (mg/m³)²

* The bigger value of: "Repeatability standard deviation at span" or
 "Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (*u_c*)

$$u_c = \sqrt{\sum (u_{max,i})^2}$$

$$U = u_c * k = u_c * 1.96$$

1.07 mg/m³
 2.10 mg/m³

Total expanded uncertainty

Relative total expanded uncertainty

U in % of the ELV mg/m³	21.0
U in % of the ELV mg/m³	30.0
U in % of the ELV mg/m³	22.5

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.1

Device data

Customer Identification	SIEMENS
Serial number	c/o Impianto Di Combustione "Termitica Colleferro"
	11220143
Date	2012-05-02

Measuring system

MCS100FT

Input values

Component	Certification range	Emissions limit value	Confidence interval
CO	75,00 mg/m ³	75,00 mg/m ³	10 %
CO ₂	25,00 Vol%	25,00 Vol%	20 %
NO	200,00 mg/m ³	200,00 ** mg/m ³	20 %
NO ₂	100,00 mg/m ³	100,00 mg/m ³	20 %
N ₂ O	50,00 mg/m ³	50,00 mg/m ³	20 %
SO ₂	76,00 mg/m ³	76,00 mg/m ³	20 %
HCl	15,00 mg/m ³	10,00 mg/m ³	40 %
HF	3,00 mg/m ³	1,00 mg/m ³	40 %
NH ₃	10,00 mg/m ³	10,00 mg/m ³	40 %
H ₂ O	40,00 Vol%	40,00 Vol%	40 %
CH ₄	50,00 mg/m ³	20,00 mg/m ³	20 %
Corg	15,00 mg/m ³	10,00 mg/m ³	30 %
O ₂	21,00 Vol%	21,00 Vol%	20 %

* For this measuring component no emission limit values and confidence intervals are defined. Therefore full scale values and exemplary confidence intervals are used here.

** The emissions limit value for NO_x is given as NO₂-concentration, therefore the value as NO-concentration is decreased by the factor 1,63.

Interferent	Concentration	Interferent	Concentration
Oxygen (O ₂)	3,00 Vol%	Ammonia (NH ₃)	20,00 mg/m ³
Oxygen (O ₂)	21,00 Vol%	Sulfur dioxide (SO ₂)	200,00 mg/m ³
Water (H ₂ O)	30,00 Vol%	Sulfur dioxide (SO ₂)	1000,00 mg/m ³
Carbon monoxide (CO)	300,00 mg/m ³	Hydrogen chloride (HCl)	50,00 mg/m ³
Carbon dioxide (CO ₂)	15,00 Vol%	Hydrogen chloride (HCl)	200,00 mg/m ³
Methane (CH ₄)	50,00 mg/m ³		
Dinitrogen oxide (N ₂ O)	20,00 mg/m ³		
Dinitrogen oxide (N ₂ O)	100,00 mg/m ³		
Nitrogen monoxide (NO)	300,00 mg/m ³		
Nitrogen dioxide (NO ₂)	30,00 mg/m ³		

Required quality of the measurement

Requirement to response time

25

%

**

Requirement of the legislation, the customer or authority

Averaging time of measured values

60

min

** Possible values are 25% for dynamic (standard) or 10 % for highly dynamic processes (EN ISO 14956, 7.2)

Summary of the results

Component	Response time	SI(AMS) values		Quality of the measurement
		Zero point	Span point	
CO	Requirements fulfilled	1,6054	2,2209	Requirements fulfilled
CO ₂	Requirements fulfilled	0,6642	0,8039	Requirements fulfilled
NO	Requirements fulfilled	4,8824	6,8904	Requirements fulfilled
NO ₂	Requirements fulfilled	3,5887	3,7303	Requirements fulfilled
N ₂ O	Requirements fulfilled	1,0101	1,3420	Requirements fulfilled
SO ₂	Requirements fulfilled	1,8490	2,8310	Requirements fulfilled
HCl	Requirements fulfilled	0,5535	0,6030	Requirements fulfilled
HF	Requirements fulfilled	0,1481	0,1281	Requirements fulfilled
NH ₃	Requirements fulfilled	0,3362	0,3091	Requirements fulfilled
H ₂ O	Requirements fulfilled	0,9910	1,0581	Requirements fulfilled
CH ₄	Requirements fulfilled	1,2035	1,5330	Requirements fulfilled
Corg	Requirements fulfilled	0,7145	0,6893	Requirements fulfilled
O ₂	Requirements fulfilled	0,2851	0,3257	Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 6.1

Device data

Customer Identification	SIEMENS	Date	2012-05-02
Serial number	c/o Impianto Di Combustione "Termica Colleferro" 11220143	Component	CO
Measuring system	MCS100FT		

Input values

Certification range	75 mg/m³	Requirement to response time	25 %
Emissions limit value	75 mg/m³	Averaging time of measured values	60 min
Confidence interval	10 %		

General Information

Maintenance Interval	6 months	Detection limit	0,32 mg/m³
----------------------	----------	-----------------	------------

Required performance regarding dynamic operating conditions

Measured response time	2,27 min	
Requirement to response time	15,00 min	25% of the averaging time of 60 min

Result

Requirements fulfilled

Calculation of the expanded uncertainty

Interferent	Zero point	Span point
3 Vol% Oxygen (O2)	0,00 mg/m³	0,00 mg/m³
21 Vol% Oxygen (O2)	0,00 mg/m³	0,00 mg/m³
30 Vol% Water (H2O)	0,00 mg/m³	0,75 mg/m³
300 mg/m³ Carbon monoxide (CO)	1,20 mg/m³	0,00 mg/m³
15 Vol% Carbon dioxide (CO2)	0,00 mg/m³	0,00 mg/m³
50 mg/m³ Methane (CH4)	0,00 mg/m³	0,00 mg/m³
20 mg/m³ Dinitrogen oxide (N2O)	0,00 mg/m³	0,00 mg/m³
100 mg/m³ Dinitrogen oxide (N2O)	-1,35 mg/m³	-2,63 mg/m³
300 mg/m³ Nitrogen monoxide (NO)	0,00 mg/m³	0,00 mg/m³
30 mg/m³ Nitrogen dioxide (NO2)	0,00 mg/m³	0,00 mg/m³
20 mg/m³ Ammonia (NH3)	0,00 mg/m³	0,53 mg/m³
200 mg/m³ Sulfur dioxide (SO2)	0,00 mg/m³	0,00 mg/m³
1000 mg/m³ Sulfur dioxide (SO2)	0,00 mg/m³	0,00 mg/m³
50 mg/m³ Hydrogen chloride (HCl)	0,00 mg/m³	0,00 mg/m³
200 mg/m³ Hydrogen chloride (HCl)	0,00 mg/m³	0,00 mg/m³

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

1,20 mg/m³
-1,35 mg/m³

1,28 mg/m³
-2,63 mg/m³

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.1

Device data

Customer Identification	SIEMENS	Date	2012-05-02
Serial number	c/o Impianto Di Combustione "Termica Colleferro" 11220143	Component	CO
Measuring system	MCS100FT		

Influences of the process characteristics

Process characteristics

- Lack-of-fit (Linearity)
- Zero drift from the field test
- Span drift from the field test
- Influence of ambient temperature at span point
- Influence of sample gas pressure
- Influence of sample gas flow
- Influence of voltage
- Cross-sensitivity
- Repeatability at span point
- Standard deviation from paired measurements under field conditions
- Uncertainty of provided reference material
- Misalignment
- Conversion rate of AMS for measurement of NOx
- Changes of response factors

Largest difference according to type approval

Zero point	Span point
1,28 mg/m³	1,28 mg/m³
1,13 mg/m³	0,00 mg/m³
0,00 mg/m³	-1,28 mg/m³
0,15 mg/m³	-1,28 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	-0,08 mg/m³
0,15 mg/m³	0,23 mg/m³
-1,35 mg/m³	-2,63 mg/m³
0,16 mg/m³	0,22 mg/m³
0,85 mg/m³	0,85 mg/m³
1,60 mg/m³	1,60 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	0,00 mg/m³

Process characteristics

- Lack-of-fit (Linearity)
- Zero drift from the field test
- Span drift from the field test
- Influence of ambient temperature at span point
- Influence of sample gas pressure
- Influence of sample gas flow
- Influence of voltage
- Cross-sensitivity
- Repeatability at span point
- Standard deviation from paired measurements under field conditions
- Uncertainty of provided reference material
- Misalignment
- Conversion rate of AMS for measurement of NOx
- Changes of response factors

Standard uncertainty

Zero point	Span point
0,7361 mg/m³	0,7361 mg/m³
0,6495 mg/m³	0,0000 mg/m³
0,0000 mg/m³	-0,7361 mg/m³
0,0866 mg/m³	-0,7361 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	-0,0433 mg/m³
0,0866 mg/m³	0,1299 mg/m³
-0,7794 mg/m³	-1,5155 mg/m³
0,0924 mg/m³	0,1270 mg/m³
0,4909 mg/m³	0,4909 mg/m³
0,8660 mg/m³	0,8660 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	0,0000 mg/m³

Calculation of the combined standard uncertainties

Combined standard uncertainty	s(AMS) values	Zero point	Span point
		1,6054 mg/m³	2,2209 mg/m³

Verification of compliance with the requirements

Combined standard uncertainty	2,31 mg/m³	according to EN 15267-3
Expanded uncertainty	4,54 mg/m³	according to EN 15267-3
Relatively expanded uncertainty	6,05 %	of the emissions limit value of 75 mg/m³
Allowed expanded uncertainty	10,00 %	of the emissions limit value of 75 mg/m³
Allowed expanded uncertainty	7,50 mg/m³	

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 6.1

Device data

Customer Identification	SIEMENS	Date	2012-05-02
Serial number	c/o Impianto Di Combustione "Termica Colleferro" 11220143	Component	CO
Measuring system	MCS100FT		

Influences of the process characteristics

Process characteristics

- Lack-of-fit (Linearity)
- Zero drift from the field test
- Span drift from the field test
- Influence of ambient temperature at span point
- Influence of sample gas pressure
- Influence of sample gas flow
- Influence of voltage
- Cross-sensitivity
- Repeatability at span point
- Standard deviation from paired measurements under field conditions
- Uncertainty of provided reference material
- Misalignment
- Conversion rate of AMS for measurement of NOx
- Changes of response factors

Largest difference according to type approval

Zero point	Span point
1,28 mg/m³	1,28 mg/m³
1,13 mg/m³	0,00 mg/m³
0,00 mg/m³	-1,28 mg/m³
0,15 mg/m³	-1,28 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	-0,08 mg/m³
0,15 mg/m³	0,23 mg/m³
-1,35 mg/m³	-2,63 mg/m³
0,16 mg/m³	0,22 mg/m³
0,85 mg/m³	0,85 mg/m³
1,50 mg/m³	1,50 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	0,00 mg/m³

Process characteristics

- Lack-of-fit (Linearity)
- Zero drift from the field test
- Span drift from the field test
- Influence of ambient temperature at span point
- Influence of sample gas pressure
- Influence of sample gas flow
- Influence of voltage
- Cross-sensitivity
- Repeatability at span point
- Standard deviation from paired measurements under field conditions
- Uncertainty of provided reference material
- Misalignment
- Conversion rate of AMS for measurement of NOx
- Changes of response factors

Standard uncertainty

Zero point	Span point
0,7361 mg/m³	0,7361 mg/m³
0,6495 mg/m³	0,0000 mg/m³
0,0000 mg/m³	-0,7361 mg/m³
0,0866 mg/m³	-0,7361 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	-0,0433 mg/m³
0,0866 mg/m³	0,1299 mg/m³
-0,7794 mg/m³	-1,5155 mg/m³
0,0924 mg/m³	0,1270 mg/m³
0,4909 mg/m³	0,4909 mg/m³
0,8660 mg/m³	0,8660 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	0,0000 mg/m³

Calculation of the combined standard uncertainties

Combined standard uncertainty	s(AMS) values	Zero point	Span point
		1,6054 mg/m³	2,2209 mg/m³

Verification of compliance with the requirements

Combined standard uncertainty	2,31	mg/m³	according to EN 15267-3
Expanded uncertainty	4,54	mg/m³	according to EN 15267-3
Relative expanded uncertainty	6,05	%	of the emissions limit value of 75 mg/m³
Allowed expanded uncertainty	10,00	%	of the emissions limit value of 75 mg/m³
Allowed expanded uncertainty	7,50	mg/m³	

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 6.1

Device data

Customer identification	SIEMENS	Date	2012-05-02
Serial number	c/o Impianto DI Combustione "Termica Colleferro"	11220143	
Measuring system	MCS100FT	Component	CO2

Input values

Certification range	25 Vol%	Requirement to response time	25 %
Measuring range	25 Vol%	Averaging time of measured values	60 min
Confidence interval	20 %		

Attention: The 2001/80/EC and 2000/70/EC gives no requirements for these components.

General information

Maintenance interval	3 months	Detection limit	0,06 Vol%
----------------------	----------	-----------------	-----------

Required performance regarding dynamic operating conditions

Measured response time	2,87 min	
Requirement to response time	15,00 min	25% of the averaging time of 60 min

Result

Requirements fulfilled

Calculation of the expanded uncertainty

Interferent	Zero point	Span point
3 Vol% Oxygen (O2)	0,00 Vol%	0,00 Vol%
21 Vol% Oxygen (O2)	0,00 Vol%	0,00 Vol%
30 Vol% Water (H2O)	-0,33 Vol%	0,48 Vol%
300 mg/m³ Carbon monoxide (CO)	0,00 Vol%	0,00 Vol%
15 Vol% Carbon dioxide (CO2)	0,00 Vol%	0,00 Vol%
50 mg/m³ Methane (CH4)	0,00 Vol%	0,00 Vol%
20 mg/m³ Dinitrogen oxide (N2O)	0,00 Vol%	-0,35 Vol%
100 mg/m³ Dinitrogen oxide (N2O)	-0,40 Vol%	0,00 Vol%
300 mg/m³ Nitrogen monoxide (NO)	0,00 Vol%	0,00 Vol%
30 mg/m³ Nitrogen dioxide (NO2)	0,00 Vol%	0,00 Vol%
20 mg/m³ Ammonia (NH3)	0,00 Vol%	-0,35 Vol%
200 mg/m³ Sulfur dioxide (SO2)	0,00 Vol%	0,00 Vol%
1000 mg/m³ Sulfur dioxide (SO2)	0,00 Vol%	0,33 Vol%
50 mg/m³ Hydrogen chloride (HCl)	0,00 Vol%	0,00 Vol%
200 mg/m³ Hydrogen chloride (HCl)	0,00 Vol%	-0,13 Vol%

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

0,00 Vol%
-0,73 Vol%

0,80 Vol%
-0,63 Vol%

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.1

Device data

Customer Identification Serial number	SIEMENS c/o Impianto DI Combustione "Termica Colleferro" 11220143	Date	2012-05-02
Measuring system	MCS100FT	Component	CO2

Influences of the process characteristics

Process characteristics

Lack-of-fit (Linearity)
Zero drift from the field test
Span drift from the field test
Influence of ambient temperature at span point
Influence of sample gas pressure
Influence of sample gas flow
Influence of voltage
Cross-sensitivity
Repeatability at span point
Standard deviation from paired measurements under field conditions
Uncertainty of provided reference material
Misalignment
Conversion rate of AMS for measurement of NOx
Changes of response factors

Largest difference according to type approval			
	Zero point	Span point	
0,38 Vol%	0,38 Vol%		
0,53 Vol%	0,00 Vol%	0,68 Vol%	
0,00 Vol%	0,08 Vol%	0,53 Vol%	
0,00 Vol%	0,00 Vol%	0,00 Vol%	
0,00 Vol%	0,00 Vol%	-0,03 Vol%	
-0,03 Vol%	-0,73 Vol%	0,10 Vol%	
-0,73 Vol%	0,03 Vol%	-0,83 Vol%	
0,03 Vol%	0,35 Vol%	0,05 Vol%	
0,35 Vol%	0,50 Vol%	0,35 Vol%	
0,50 Vol%	0,00 Vol%	0,00 Vol%	
0,00 Vol%	0,00 Vol%	0,00 Vol%	
0,00 Vol%	0,00 Vol%	0,00 Vol%	

Process characteristics

Lack-of-fit (Linearity)
Zero drift from the field test
Span drift from the field test
Influence of ambient temperature at span point
Influence of sample gas pressure
Influence of sample gas flow
Influence of voltage
Cross-sensitivity
Repeatability at span point
Standard deviation from paired measurements under field conditions
Uncertainty of provided reference material
Misalignment
Conversion rate of AMS for measurement of NOx
Changes of response factors

Standard uncertainty			
	Zero point	Span point	
$U_{\text{lof}} =$	0,2165 Vol%	0,2165 Vol%	
$U_{d,z} =$	0,3031 Vol%	0,0000 Vol%	
$U_{d,s} =$	0,0000 Vol%	0,3897 Vol%	
$U_t =$	0,0433 Vol%	0,3031 Vol%	
$U_p =$	0,0000 Vol%	0,0000 Vol%	
$U_f =$	0,0000 Vol%	-0,0144 Vol%	
$U_v =$	-0,0144 Vol%	0,0577 Vol%	
$U_i =$	-0,4186 Vol%	-0,4763 Vol%	
$U_r =$	0,0173 Vol%	0,0289 Vol%	
$U_d =$	0,2046 Vol%	0,2046 Vol%	
$U_m =$	0,2887 Vol%	0,2887 Vol%	
$U_{mb} =$	0,0000 Vol%	0,0000 Vol%	
$U_{ce} =$	0,0000 Vol%	0,0000 Vol%	
$U_{rf} =$	0,0000 Vol%	0,0000 Vol%	

Calculation of the combined standard uncertainty

Combined standard uncertainty	s(AMS) values	Zero point	Span point
		0,6642 Vol%	0,8039 Vol%

Verification of compliance with the requirements

Combined standard uncertainty	0,86	Vol%	according to EN 15267-3
Expanded uncertainty	1,68	Vol%	according to EN 15267-3
Relative expanded uncertainty	6,74	%	of the measuring range of 25 Vol%
Allowed expanded uncertainty	20,00	%	of the measuring range of 25 Vol%
Allowed expanded uncertainty	5,00	Vol%	

Result

Requirements fulfilled

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 6.1

Device data

Customer Identification Serial number	SIEMENS c/o Impianto DI Combustione "Termica Colleferro" 11220143	Date	2012-05-02
Measuring system	MCS100FT	Component	NO

Input values

Certification range	200	mg/m³	Requirement to response time	25	%
Emissions limit value	200	mg/m³	Averaging time of measured values	60	min
Confidence interval	20	%			

General information

Maintenance interval	6	months	Detection limit	0,76	mg/m³
----------------------	---	--------	-----------------	------	-------

Required performance regarding dynamic operating conditions

Measured response time	2,87	min	
Requirement to response time	15,00	min	25% of the averaging time of 60 min

Result

Requirements fulfilled

Calculation of the expanded uncertainty

Interferent

3 Vol% Oxygen (O ₂)	0,00	mg/m³
21 Vol% Oxygen (O ₂)	0,00	mg/m³
30 Vol% Water (H ₂ O)	-5,20	mg/m³
300 mg/m³ Carbon monoxide (CO)	0,00	mg/m³
15 Vol% Carbon dioxide (CO ₂)	0,00	mg/m³
50 mg/m³ Methane (CH ₄)	0,00	mg/m³
20 mg/m³ Dinitrogen oxide (N ₂ O)	0,00	mg/m³
100 mg/m³ Dinitrogen oxide (N ₂ O)	0,00	mg/m³
300 mg/m³ Nitrogen monoxide (NO)	0,00	mg/m³
30 mg/m³ Nitrogen dioxide (NO ₂)	0,00	mg/m³
20 mg/m³ Ammonia (NH ₃)	1,40	mg/m³
200 mg/m³ Sulfur dioxide (SO ₂)	0,00	mg/m³
1000 mg/m³ Sulfur dioxide (SO ₂)	0,00	mg/m³
50 mg/m³ Hydrogen chloride (HCl)	0,00	mg/m³
200 mg/m³ Hydrogen chloride (HCl)	0,00	mg/m³

Zero point

0,00	mg/m³	0,00	mg/m³
0,00	mg/m³	0,00	mg/m³
-5,20	mg/m³	3,20	mg/m³
0,00	mg/m³	0,00	mg/m³
0,00	mg/m³	-4,80	mg/m³
0,00	mg/m³	0,00	mg/m³
0,00	mg/m³	0,00	mg/m³
0,00	mg/m³	2,00	mg/m³
0,00	mg/m³	0,00	mg/m³
1,40	mg/m³	0,00	mg/m³
0,00	mg/m³	0,00	mg/m³
0,00	mg/m³	1,20	mg/m³
0,00	mg/m³	0,00	mg/m³
0,00	mg/m³	1,60	mg/m³

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

1,40	mg/m³
-5,20	mg/m³

8,00	mg/m³
-4,80	mg/m³

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.1

Device data

Customer Identification Serial number	SIEMENS c/o Impianto Di Combustione "Termica Colleferro" 11220143	Date	2012-05-02
Measuring system	MCS100FT	Component	NO

Influences of the process characteristics

Process characteristics

Lack-of-fit (Linearity)
 Zero drift from the field test
 Span drift from the field test
 Influence of ambient temperature at span point
 Influence of sample gas pressure
 Influence of sample gas flow
 Influence of voltage
 Cross-sensitivity
 Repeatability at span point
 Standard deviation from paired measurements under field conditions
 Uncertainty of provided reference material
 Misalignment
 Conversion rate of AMS for measurement of NOx
 Changes of response factors

Largest difference according to type approval			
	Zero point	Span point	
	-2,00 mg/m³	-2,00 mg/m³	
	3,60 mg/m³	0,00 mg/m³	
	0,00 mg/m³	-6,00 mg/m³	
	-0,80 mg/m³	-3,00 mg/m³	
	0,00 mg/m³	0,00 mg/m³	
	0,00 mg/m³	-0,20 mg/m³	
	0,20 mg/m³	-1,60 mg/m³	
	-5,20 mg/m³	8,00 mg/m³	
	0,38 mg/m³	0,78 mg/m³	
	3,29 mg/m³	3,29 mg/m³	
	4,00 mg/m³	4,00 mg/m³	
	0,00 mg/m³	0,00 mg/m³	
	0,00 mg/m³	0,00 mg/m³	
	0,00 mg/m³	0,00 mg/m³	

Process characteristics

Lack-of-fit (Linearity)
 Zero drift from the field test
 Span drift from the field test
 Influence of ambient temperature at span point
 Influence of sample gas pressure
 Influence of sample gas flow
 Influence of voltage
 Cross-sensitivity
 Repeatability at span point
 Standard deviation from paired measurements under field conditions
 Uncertainty of provided reference material
 Misalignment
 Conversion rate of AMS for measurement of NOx
 Changes of response factors

Standard uncertainty			
	Zero point	Span point	
U_{lf} =	-1,1547 mg/m³	-1,1547 mg/m³	
$U_{d,z}$ =	2,0785 mg/m³	0,0000 mg/m³	
$U_{d,s}$ =	0,0000 mg/m³	-3,4641 mg/m³	
U_t =	-0,4619 mg/m³	-1,7321 mg/m³	
U_p =	0,0000 mg/m³	0,0000 mg/m³	
U_r =	0,0000 mg/m³	-0,1155 mg/m³	
U_v =	0,1165 mg/m³	-0,9238 mg/m³	
U_l =	-3,0022 mg/m³	4,6188 mg/m³	
U_r =	0,2194 mg/m³	0,4503 mg/m³	
U_p =	1,9004 mg/m³	1,9004 mg/m³	
U_m =	2,3094 mg/m³	2,3094 mg/m³	
U_{mb} =	0,0000 mg/m³	0,0000 mg/m³	
U_{ca} =	0,0000 mg/m³	0,0000 mg/m³	
U_{if} =	0,0000 mg/m³	0,0000 mg/m³	

Calculation of the combined standard uncertainties

Combined standard uncertainty	s(AMS) values	Zero point	Span point
		4,8824 mg/m³	6,8904 mg/m³

Verification of compliance with the requirements

Combined standard uncertainty	7,20 mg/m³	according to EN 15267-3
Expanded uncertainty	14,11 mg/m³	according to EN 15267-3
Relative expanded uncertainty	7,05 %	of the emissions limit value of 200 mg/m³
Allowed expanded uncertainty	20,00 %	of the emissions limit value of 200 mg/m³
Allowed expanded uncertainty	40,00 mg/m³	

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.1

Device data

Customer Identification	SIEMENS	Date	2012-05-02
Serial number	c/o Impianto Di Combustione "Termica Colleferro" 11220143	Component	NO2
Measuring system	MCS100FT		

Input values

Certification range	100 mg/m³	Requirement to response time	25 %
Emissions limit value	100 mg/m³	Averaging time of measured values	60 min
Confidence interval	20 %		

General information

Maintenance interval	6 months	Detection limit	0,38 mg/m³
----------------------	----------	-----------------	------------

Required performance regarding dynamic operating conditions

Measured response time	2,92 min	
Requirement to response time	15,00 min	25% of the averaging time of 60 min

Result

Requirements fulfilled

Calculation of the expanded uncertainty

Interferent

Interferent	Zero point	Span point
3 Vol% Oxygen (O2)	0,00 mg/m³	0,00 mg/m³
21 Vol% Oxygen (O2)	0,00 mg/m³	0,00 mg/m³
30 Vol% Water (H2O)	-1,50 mg/m³	1,00 mg/m³
300 mg/m³ Carbon monoxide (CO)	0,00 mg/m³	0,00 mg/m³
15 Vol% Carbon dioxide (CO2)	0,50 mg/m³	-1,10 mg/m³
50 mg/m³ Methane (CH4)	-0,90 mg/m³	0,00 mg/m³
20 mg/m³ Dinitrogen oxide (N2O)	0,00 mg/m³	0,00 mg/m³
100 mg/m³ Dinitrogen oxide (N2O)	1,00 mg/m³	0,00 mg/m³
300 mg/m³ Nitrogen monoxide (NO)	3,20 mg/m³	3,00 mg/m³
30 mg/m³ Nitrogen dioxide (NO2)		
20 mg/m³ Ammonia (NH3)	0,00 mg/m³	-1,00 mg/m³
200 mg/m³ Sulfur dioxide (SO2)	0,00 mg/m³	0,00 mg/m³
1000 mg/m³ Sulfur dioxide (SO2)	0,00 mg/m³	0,00 mg/m³
50 mg/m³ Hydrogen chloride (HCl)	0,00 mg/m³	0,00 mg/m³
200 mg/m³ Hydrogen chloride (HCl)	0,00 mg/m³	-2,60 mg/m³

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

4,70 mg/m³
-2,40 mg/m³

4,00 mg/m³
-4,70 mg/m³

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.1

Device data

Customer identification	SIEMENS c/o Impianto Di Combustione "Termica Colleferro"	Date	2012-05-02
Serial number	11220143	Component	NO2

Influences of the process characteristics

Process characteristics

- Lack-of-fit (Linearity)
- Zero drift from the field test
- Span drift from the field test
- Influence of ambient temperature at span point
- Influence of sample gas pressure
- Influence of sample gas flow
- Influence of voltage
- Cross-sensitivity
- Repeatability at span point
- Standard deviation from paired measurements under field conditions
- Uncertainty of provided reference material
- Misalignment
- Conversion rate of AMS for measurement of NOx
- Changes of response factors

Largest difference according to type approval

Zero point	Span point
2,00 mg/m³	2,00 mg/m³
-2,30 mg/m³	0,00 mg/m³
0,00 mg/m³	3,00 mg/m³
1,30 mg/m³	0,90 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	-0,10 mg/m³
0,30 mg/m³	-0,60 mg/m³
4,70 mg/m³	-4,70 mg/m³
0,19 mg/m³	0,47 mg/m³
1,21 mg/m³	1,21 mg/m³
2,00 mg/m³	2,00 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	0,00 mg/m³

Process characteristics

- Lack-of-fit (Linearity)
- Zero drift from the field test
- Span drift from the field test
- Influence of ambient temperature at span point
- Influence of sample gas pressure
- Influence of sample gas flow
- Influence of voltage
- Cross-sensitivity
- Repeatability at span point
- Standard deviation from paired measurements under field conditions
- Uncertainty of provided reference material
- Misalignment
- Conversion rate of AMS for measurement of NOx
- Changes of response factors

Standard uncertainty

Zero point	Span point
1,1547 mg/m³	1,1547 mg/m³
-1,3279 mg/m³	0,0000 mg/m³
0,0000 mg/m³	1,7321 mg/m³
0,7506 mg/m³	0,5196 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	-0,0577 mg/m³
0,1732 mg/m³	-0,3464 mg/m³
2,7135 mg/m³	-2,7135 mg/m³
0,1097 mg/m³	0,2714 mg/m³
0,7013 mg/m³	0,7013 mg/m³
1,1547 mg/m³	1,1547 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	0,0000 mg/m³

Calculation of the combined standard uncertainties

Combined standard uncertainty	s(AMS) values	Zero point	Span point
		3,5887 mg/m³	3,7303 mg/m³

Verification of compliance with the requirements

Combined standard uncertainty	3,96 mg/m³	according to EN 15267-3
Expanded uncertainty	7,76 mg/m³	according to EN 15267-3
Relative expanded uncertainty	7,76 %	of the emissions limit value of 100 mg/m³
Allowed expanded uncertainty	20,00 %	of the emissions limit value of 100 mg/m³
Allowed expanded uncertainty	20,00 mg/m³	

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.1

Device data

Customer Identification Serial number	SIEMENS c/o Impianto Di Combustione "Termica Colleferro" 11220143	Date	2012-05-02
Measuring system	MCS100FT	Component	NH3

Input values

Certification range	10 mg/m³	Requirement to response time	25 %
Measuring range	10 mg/m³	Averaging time of measured values	60 min
Confidence interval	40 %	*	

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

General information

Maintenance interval	4 weeks	Detection limit	0,05 mg/m³
----------------------	---------	-----------------	------------

Required performance regarding dynamic operating conditions

Measured response time	3,18 min	
Requirement to response time	15,00 min	25% of the averaging time of 60 min

Result

Requirements fulfilled

Calculation of the expanded uncertainty

Interferent	Zero point	Span point
3 Vol% Oxygen (O2)	0,00 mg/m³	0,00 mg/m³
21 Vol% Oxygen (O2)	0,00 mg/m³	0,00 mg/m³
30 Vol% Water (H2O)	0,08 mg/m³	0,00 mg/m³
300 mg/m³ Carbon monoxide (CO)	0,07 mg/m³	0,00 mg/m³
15 Vol% Carbon dioxide (CO2)	0,09 mg/m³	-0,10 mg/m³
50 mg/m³ Methane (CH4)	0,12 mg/m³	-0,04 mg/m³
20 mg/m³ Dinitrogen oxide (N2O)	0,00 mg/m³	0,00 mg/m³
100 mg/m³ Dinitrogen oxide (N2O)	0,00 mg/m³	0,04 mg/m³
300 mg/m³ Nitrogen monoxide (NO)	0,13 mg/m³	-0,10 mg/m³
30 mg/m³ Nitrogen dioxide (NO2)	0,00 mg/m³	-0,05 mg/m³
20 mg/m³ Ammonia (NH3)		
200 mg/m³ Sulfur dioxide (SO2)	0,00 mg/m³	0,00 mg/m³
1000 mg/m³ Sulfur dioxide (SO2)	0,00 mg/m³	0,00 mg/m³
50 mg/m³ Hydrogen chloride (HCl)	0,00 mg/m³	0,00 mg/m³
200 mg/m³ Hydrogen chloride (HCl)	0,00 mg/m³	0,00 mg/m³

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

0,49 mg/m³
0,00 mg/m³

0,04 mg/m³
-0,29 mg/m³

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 6.1

Device data

Customer Identification Serial number	SIEMENS c/o Impianto DI Combustione "Termica Colleferro" 11220143	Date	2012-05-02
Measuring system	MCS100FT	Component	NH3

Influences of the process characteristics

Process characteristics

Lack-of-fit (Linearity)
 Zero drift from the field test
 Span drift from the field test
 Influence of ambient temperature at span point
 Influence of sample gas pressure
 Influence of sample gas flow
 Influence of voltage
 Cross-sensitivity
 Repeatability at span point
 Standard deviation from paired measurements under field conditions
 Uncertainty of provided reference material
 Misalignment
 Conversion rate of AMS for measurement of NOx
 Changes of response factors

Largest difference according to type approval	
Zero point	Span point
0,18 mg/m³	0,18 mg/m³
0,05 mg/m³	0,00 mg/m³
0,00 mg/m³	0,29 mg/m³
-0,05 mg/m³	-0,14 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	-0,01 mg/m³
-0,05 mg/m³	0,07 mg/m³
0,49 mg/m³	-0,29 mg/m³
0,02 mg/m³	0,07 mg/m³
0,14 mg/m³	0,14 mg/m³
0,20 mg/m³	0,20 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	0,00 mg/m³
0,00 mg/m³	0,00 mg/m³

Process characteristics

Lack-of-fit (Linearity)
 Zero drift from the field test
 Span drift from the field test
 Influence of ambient temperature at span point
 Influence of sample gas pressure
 Influence of sample gas flow
 Influence of voltage
 Cross-sensitivity
 Repeatability at span point
 Standard deviation from paired measurements under field conditions
 Uncertainty of provided reference material
 Misalignment
 Conversion rate of AMS for measurement of NOx
 Changes of response factors

Standard uncertainty	
Zero point	Span point
0,1039 mg/m³	0,1039 mg/m³
0,0289 mg/m³	0,0000 mg/m³
0,0000 mg/m³	0,1697 mg/m³
-0,0289 mg/m³	-0,0808 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	-0,0058 mg/m³
-0,0289 mg/m³	0,0404 mg/m³
0,2829 mg/m³	-0,1674 mg/m³
0,0115 mg/m³	0,0404 mg/m³
0,0796 mg/m³	0,0796 mg/m³
0,1155 mg/m³	0,1155 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	0,0000 mg/m³
0,0000 mg/m³	0,0000 mg/m³

Calculation of the combined standard uncertainties

Combined standard uncertainty	s(AMS) values	Zero point	Span point
		0,3362 mg/m³	0,3091 mg/m³

Verification of compliance with the requirements

Combined standard uncertainty	0,39	mg/m³	according to EN 15267-3
Expanded uncertainty	0,75	mg/m³	according to EN 15267-3
Relative expanded uncertainty	7,55	%	of the measuring range of 10 mg/m³
Allowed expanded uncertainty	40,00	%	of the measuring range of 10 mg/m³
Allowed expanded uncertainty	4,00	mg/m³	

Result

Requirements fulfilled

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.1

Device data

Customer Identification	SIEMENS		
Serial number	c/o Impianto Di Combustione "Termica Colleferro"	11220143	Date
Measuring system	MCS100FT	Component	H2O

Input values

Certification range	40 Vol%	Requirement to response time	25 %
Measuring range	40 Vol%	Averaging time of measured values	60 min
Confidence interval	40 %		

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

General information

Maintenance Interval	6 months	Detection limit	0,04 Vol%
----------------------	----------	-----------------	-----------

Required performance regarding dynamic operating conditions

Measured response time	3,00 min	
Requirement to response time	15,00 min	25% of the averaging time of 60 min

Result

Requirements fulfilled

Calculation of the expanded uncertainty

Interferent	Zero point	Span point
3 Vol% Oxygen (O2)	0,00 Vol%	0,00 Vol%
21 Vol% Oxygen (O2)	0,00 Vol%	0,00 Vol%
30 Vol% Water (H2O)		
300 mg/m³ Carbon monoxide (CO)	0,80 Vol%	0,76 Vol%
15 Vol% Carbon dioxide (CO2)	0,00 Vol%	0,00 Vol%
50 mg/m³ Methane (CH4)	-0,20 Vol%	-0,36 Vol%
20 mg/m³ Dinitrogen oxide (N2O)	0,00 Vol%	0,00 Vol%
100 mg/m³ Dinitrogen oxide (N2O)	0,00 Vol%	0,00 Vol%
300 mg/m³ Nitrogen monoxide (NO)	0,00 Vol%	0,00 Vol%
30 mg/m³ Nitrogen dioxide (NO2)	0,00 Vol%	0,00 Vol%
20 mg/m³ Ammonia (NH3)	0,00 Vol%	-0,20 Vol%
200 mg/m³ Sulfur dioxide (SO2)	0,00 Vol%	0,00 Vol%
1000 mg/m³ Sulfur dioxide (SO2)	0,00 Vol%	0,00 Vol%
50 mg/m³ Hydrogen chloride (HCl)	0,00 Vol%	0,00 Vol%
200 mg/m³ Hydrogen chloride (HCl)	0,00 Vol%	-0,20 Vol%

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

0,80 Vol%
-0,20 Vol%

0,76 Vol%
-0,76 Vol%

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 6.1

Device data

Customer Identification	SIEMENS	Date	2012-05-02
Serial number	c/o Impianto Di Combustione "Termica Colleferro" 11220143	Component	H2O
Measuring system	MCS100FT		

Influences of the process characteristics

Process characteristics

- Lack-of-fit (Linearity)
- Zero drift from the field test
- Span drift from the field test
- Influence of ambient temperature at span point
- Influence of sample gas pressure
- Influence of sample gas flow
- Influence of voltage
- Cross-sensitivity
- Repeatability at span point
- Standard deviation from paired measurements under field conditions
- Uncertainty of provided reference material
- Misalignment
- Conversion rate of AMS for measurement of NOx
- Changes of response factors

Largest difference according to type approval

Zero point	Span point
0,64 Vol%	0,64 Vol%
-1,04 Vol%	0,00 Vol%
0,00 Vol%	1,16 Vol%
0,12 Vol%	0,48 Vol%
0,00 Vol%	0,00 Vol%
0,00 Vol%	-0,04 Vol%
0,08 Vol%	0,00 Vol%
0,80 Vol%	0,76 Vol%
0,02 Vol%	0,06 Vol%
0,39 Vol%	0,39 Vol%
0,80 Vol%	0,80 Vol%
0,00 Vol%	0,00 Vol%
0,00 Vol%	0,00 Vol%
0,00 Vol%	0,00 Vol%

Process characteristics

- Lack-of-fit (Linearity)
- Zero drift from the field test
- Span drift from the field test
- Influence of ambient temperature at span point
- Influence of sample gas pressure
- Influence of sample gas flow
- Influence of voltage
- Cross-sensitivity
- Repeatability at span point
- Standard deviation from paired measurements under field conditions
- Uncertainty of provided reference material
- Misalignment
- Conversion rate of AMS for measurement of NOx
- Changes of response factors

Standard uncertainty

	Zero point	Span point
U_{tot} =	0,3695 Vol%	0,3695 Vol%
$U_{d,z}$ =	-0,6004 Vol%	0,0000 Vol%
$U_{d,s}$ =	0,0000 Vol%	0,6697 Vol%
U_t =	0,0693 Vol%	0,2771 Vol%
U_p =	0,0000 Vol%	0,0000 Vol%
U_f =	0,0000 Vol%	-0,0231 Vol%
U_v =	0,0462 Vol%	0,0000 Vol%
U_i =	0,4619 Vol%	0,4388 Vol%
U_r =	0,0115 Vol%	0,0346 Vol%
U_D =	0,2266 Vol%	0,2266 Vol%
U_m =	0,4619 Vol%	0,4619 Vol%
U_{mb} =	0,0000 Vol%	0,0000 Vol%
U_{co} =	0,0000 Vol%	0,0000 Vol%
U_{rf} =	0,0000 Vol%	0,0000 Vol%

Calculation of the combined standard uncertainties

Combined standard uncertainty	s(AMS) values	Zero point	Span point
		0,9910 Vol%	1,0581 Vol%

Verification of compliance with the requirements

Combined standard uncertainty	1,23	Vol%	according to EN 15267-3
Expanded uncertainty	2,40	Vol%	according to EN 15267-3
Relative expanded uncertainty	6,01	%	of the measuring range of 40 Vol%
Allowed expanded uncertainty	40,00	%	of the measuring range of 40 Vol%
Allowed expanded uncertainty	16,00	Vol%	

Result

Requirements fulfilled

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 6.1

Device data

Customer identification	SIEMENS		
Serial number	c/o Impianto Di Combustione "Termica Colleferro"	11220143	Date
Measuring system	MCS100FT	Component	O2

Input values

Certification range	21 Vol%	Requirement to response time	25 %
Measuring range	21 Vol%	Averaging time of measured values	60 min
Confidence interval	20 %		

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

General information

Maintenance interval	4 weeks	Detection limit	0,03 Vol%
----------------------	---------	-----------------	-----------

Required performance regarding dynamic operating conditions

Measured response time	2,27 min	
Requirement to response time	15,00 min	25% of the averaging time of 60 min

Result

Requirements fulfilled

Calculation of the expanded uncertainty

Interferent	Zero point	Span point
3 Vol% Oxygen (O2)	0,00 Vol%	0,00 Vol%
21 Vol% Oxygen (O2)	0,00 Vol%	0,00 Vol%
30 Vol% Water (H2O)	0,00 Vol%	0,00 Vol%
300 mg/m³ Carbon monoxide (CO)	0,00 Vol%	0,00 Vol%
15 Vol% Carbon dioxide (CO2)	0,00 Vol%	0,00 Vol%
50 mg/m³ Methane (CH4)	0,00 Vol%	0,00 Vol%
20 mg/m³ Dinitrogen oxide (N2O)	0,00 Vol%	0,00 Vol%
100 mg/m³ Dinitrogen oxide (N2O)	0,00 Vol%	0,00 Vol%
300 mg/m³ Nitrogen monoxide (NO)	0,00 Vol%	0,00 Vol%
30 mg/m³ Nitrogen dioxide (NO2)	0,00 Vol%	0,00 Vol%
20 mg/m³ Ammonia (NH3)	0,00 Vol%	0,00 Vol%
200 mg/m³ Sulfur dioxide (SO2)	0,00 Vol%	0,00 Vol%
1000 mg/m³ Sulfur dioxide (SO2)	0,00 Vol%	0,00 Vol%
50 mg/m³ Hydrogen chloride (HCl)	0,00 Vol%	0,00 Vol%
200 mg/m³ Hydrogen chloride (HCl)	0,00 Vol%	0,00 Vol%

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

0,00 Vol%
0,00 Vol%

0,00 Vol%
0,00 Vol%

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.1

Device data

Customer Identification Serial number	SIEMENS c/o Impianto Di Combustione "Termica Colleferro" 11220143	Date	2012-05-02
Measuring system	MCS100FT	Component	O2

Influences of the process characteristics

Process characteristics

Lack-of-fit (Linearity)
Zero drift from the field test
Span drift from the field test
Influence of ambient temperature at span point
Influence of sample gas pressure
Influence of sample gas flow
Influence of voltage
Cross-sensitivity
Repeatability at span point
Standard deviation from paired measurements under field conditions
Uncertainty of provided reference material
Misalignment
Conversion rate of AMS for measurement of NOx
Changes of response factors

Largest difference according to type approval		
	Zero point	Span point
	0,16 Vol%	0,16 Vol%
	0,18 Vol%	0,00 Vol%
	0,00 Vol%	-0,20 Vol%
	0,02 Vol%	0,24 Vol%
	0,00 Vol%	0,00 Vol%
	-0,02 Vol%	0,01 Vol%
	0,01 Vol%	-0,10 Vol%
	0,00 Vol%	0,00 Vol%
	0,01 Vol%	0,01 Vol%
	0,09 Vol%	0,09 Vol%
	0,42 Vol%	0,42 Vol%
	0,00 Vol%	0,00 Vol%
	0,00 Vol%	0,00 Vol%
	0,00 Vol%	0,00 Vol%

Process characteristics

Lack-of-fit (Linearity)
Zero drift from the field test
Span drift from the field test
Influence of ambient temperature at span point
Influence of sample gas pressure
Influence of sample gas flow
Influence of voltage
Cross-sensitivity
Repeatability at span point
Standard deviation from paired measurements under field conditions
Uncertainty of provided reference material
Misalignment
Conversion rate of AMS for measurement of NOx
Changes of response factors

Standard uncertainty		
	Zero point	Span point
U_{lof} =	0,0924 Vol%	0,0924 Vol%
$U_{d,z}$ =	0,1039 Vol%	0,0000 Vol%
$U_{d,s}$ =	0,0000 Vol%	-0,1155 Vol%
U_t =	0,0115 Vol%	0,1386 Vol%
U_p =	0,0000 Vol%	0,0000 Vol%
U_f =	-0,0115 Vol%	0,0058 Vol%
U_y =	0,0058 Vol%	-0,0577 Vol%
U_i =	0,0000 Vol%	0,0000 Vol%
U_r =	0,0058 Vol%	0,0058 Vol%
U_D =	0,0533 Vol%	0,0533 Vol%
U_m =	0,2425 Vol%	0,2425 Vol%
U_{mb} =	0,0000 Vol%	0,0000 Vol%
U_{cv} =	0,0000 Vol%	0,0000 Vol%
U_{rf} =	0,0000 Vol%	0,0000 Vol%

Calculation of the combined standard uncertainties

Combined standard uncertainty	s(AMS) values	Zero point	Span point
		0,2851 Vol%	0,3257 Vol%


Verification of compliance with the requirements

Combined standard uncertainty	0,34	Vol%	according to EN 15267-3
Expanded uncertainty	0,67	Vol%	according to EN 15267-3
Relative expanded uncertainty	3,19	%	of the measuring range of 21 Vol%
Allowed expanded uncertainty	20,00	%	of the measuring range of 21 Vol%
Allowed expanded uncertainty	4,20	Vol%	

Result

Requirements fulfilled

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.





Certificate

TÜV Süd Industrie Service GmbH

Laboratory for Environmental Services
(Laboratorium Umwelt Service)

accredited according DIN EN ISO/IEC 17025 DAP-PL-2885.99

Oxymat 6E,F 7MB20

Gas Analyser for O₂

Report Nr. 24019084 (February 1999)

Manufacturer:
Siemens AG, Karlsruhe, Germany

TÜV Süd Industrie Service GmbH is herewith certifying that the analyser Oxymat 6E,F 7MB20 for O₂ has the following expanded uncertainty (calculated according DIN EN ISO 14956, Jan. 2003 and prEN 15267-3, August 2005):

Component	C _{test} Vol.-%	Range of measurement Vol.-%	Expanded Uncertainty according EN ISO 14956
Oxygen, O ₂	11	0-25	0,49 Vol.-% (2 % of range of measurement)

The analyser can be used in combination with other tested measuring systems which fulfil QAL 1 of EN 14181.

The calculation according DIN EN ISO 14956 was performed on the basis of the results of the investigations of report 24019084 (February 1999) for the German suitability test. The following performance characteristics were regarded: Response time; lower detection limit; lack of fit; instability/ drift; repeatability; sensitivity to ambient temperature, ambient pressure, voltage supply and gas flow; sample losses, selectivity/ interfering components; uncertainty of calibration gas

Munich, January 2006

Dr. D. Fiederer
Dr. D. Fiederer

Laboratorium Umwelt Service, TÜV Süd Industrie Service GmbH, IS-US3-MUC,
Westendstrasse 199, D-80686 München

Dr. A. Brandl
Dr. A. Brandl

TÜV Süddeutschland

TÜV Ecoplan Umwelt GmbH

Prozeß-Gasanalysatoren ULTRAMAT 6 und OXYMAT 6

Bericht-Nr. 24019084

Hersteller:

SIEMENS AG, Karlsruhe, Deutschland

Die TÜV Ecoplan Umwelt GmbH, Unternehmensgruppe TÜV Süddeutschland bescheinigt hiermit für die Prozeßgasanalysatoren ULTRAMAT 6 und OXYMAT 6, daß folgende Meßergebnisse bei der Emissionsüberwachung einer Müllverbrennungsanlage erreicht wurden und die Anforderungen gemäß 13. BlmSchV, 17. BlmSchV und TA-Luft erfüllt werden.

Kleinste geprüfte Meßbereiche:

CO	0 ... 50 mg/m ³	NO	0 ... 100 mg/m ³
SO ₂	0 ... 75 mg/m ³	O ₂	0 ... 5 / 25 Vol.-%

Verfügbarkeit:

> 99% über einen Zeitraum von 3 Monaten für zwei unabhängige Systeme mit Meßgasaufbereitung.

Drift:

Mit automatischer Kalibrierung des Nullpunkts und des Endpunkts mit Kalibriergas (Intervall wöchentlich):

Nullpunkt-Drift < 2% der Meßspanne pro Jahr

Endpunkt-Drift < 2% vom Sollwert pro Jahr

Querempfindlichkeit:

Die Summe aller Querempfindlichkeiten der genannten Komponenten beträgt gegenüber SO₂, NO, O₂, NH₃, NO₂, CH₄, N₂O, CO und H₂O mit typischen Rauchgaskonzentrationen < 4% des Meßbereichs.

Nachweisgrenze:

CO	≤ 1% des MBE	NO	≤ 0,8% des MBE
SO ₂	≤ 0,7% des MBE	O ₂	≤ 0,02 Vol.-% O ₂

TÜV Ecoplan Umwelt GmbH
Unternehmensgruppe TÜV Süddeutschland
Westendstraße 199 D-80686 München
Tel. (089) 5791-2821 Fax (089) 5791-2822

Dr. A. Brandl, München, September 1999, Dr. A. Witthuhn,
TÜV Ecoplan Umwelt GmbH, Westendstrasse 199, D-80686 München

Manufacturer's Declaration of Conformity

for Automated Measuring Systems (AMS)

according to the requirements of EN 14956 and
QAL 1 according to EN 14181

SIEMENS AG A&D PI 2
76181 Karlsruhe, Germany


declares that the product

ULTRAMAT 23

7 MB 2335, 7 MB 2337

CO 0-150 mg/m³

complies with the requirements of QAL 1 according to the international
standards EN 14956 and EN 14181 for the following specified
operating conditions:



Dr. Frank Diedrich
General Manager
A&D PI 2
Siemens AG
Datum: 15.06.2005



i.v.
Peter Berghäuser
R&D Manager
A&D PI 2 RD
Siemens AG
Datum: 15.06.2005

Manufacturer's Declaration of Conformity for Automated Measuring Systems (AMS)
according to the requirements of EN 14956 and QAL 1 according to EN 14181

Specification of the Automated Measuring System

Gas analyzer
Order information
Measured component
Smallest TÜV certified measuring range



Range of Applications

Test gas concentration /		
Emission limit value (daily average)	50	mg/m³
Ambient pressure range	990 ... 1010	hPa
Ambient temperature range	20 ... 35	°C
Flow range	30 ... 90	l/h
Voltage range	190 ... 250	V

Determined Standard Uncertainties referred to Daily Average Limit Value

Non-linearity	0,147	mg/m³
Drift	0,361	mg/m³
Pressure dependence	0,000	mg/m³
Ambient temperature dependence	-0,268	mg/m³
Flow dependence	0,000	mg/m³
Voltage dependence	0,000	mg/m³
Uncertainty of test gas	0,577	mg/m³
Leakage during sampling and sample transport	0,000	mg/m³
Reference measuring method	0,323	mg/m³
Reproducibility standard deviation	0,188	mg/m³
Selectivity (cross interference):		
O2	0,520	mg/m³
CO	0,000	mg/m³
CO2	2,078	mg/m³
CH4	0,000	mg/m³
N2O	0,058	mg/m³
NO	0,000	mg/m³
NO2	0,297	mg/m³
NH3	0,000	mg/m³
SO2 (coal firing without desulfurization)	0,000	mg/m³
HCl (coal firing)	0,000	mg/m³
H2O (sample conditioning with cooler)	0,000	mg/m³

Result

Target value	< 5	mg/m³	according to 13. BImSchV
Result 95% confidence interval	4,64	mg/m³	equivalent to s _{AMS} acc. to EN 14181
<i>equals the extended measurement uncertainty</i>			
Combined standard uncertainty	2,32	mg/m³	95% confidence interval met

Response Time

Target response time	< 200	s	Requirement fulfilled
Measured response time	67	s	

Data base on: suitability test Ultramat 23 7MB2337, December 2004
Report-No: 427899, TÜV Industrie Service GmbH, TÜV Süd Gruppe

Manufacturer's Declaration of Conformity

for Automated Measuring Systems (AMS)

according to the requirements of EN 14956 and
QAL 1 according to EN 14181

SIEMENS AG A&D PI 2
76181 Karlsruhe, Germany


declares that the product

ULTRAMAT 23

7 MB 2335, 7 MB 2337

NO 0-100 mg/m³

complies with the requirements of QAL 1 according to the international
standards EN 14956 and EN 14181 for the following specified
operating conditions:




Dr. Frank Diedrich
General Manager
A&D PI 2
Siemens AG
Datum: 15.06.2005



Peter Berghäuser
R&D Manager
A&D PI 2 RD
Siemens AG
Datum: 15.06.2005

Manufacturer's Declaration of Conformity for Automated Measuring Systems (AMS)
according to the requirements of EN 14956 and QAL 1 according to EN 14181**Specification of the Automated Measuring System**

Gas analyzer
Order information
Measured component
Smallest TÜV certified measuring range

**Range of Applications**

Test gas concentration /		
Emission limit value (daily average)	33	mg/m³
Ambient pressure range	990 ... 1010	hPa
Ambient temperature range	20 ... 35	°C
Flow range	30 ... 90	l/h
Voltage range	190 ... 250	V

Determined Standard Uncertainties referred to Daily Average Limit Value

Non-linearity	0,110	mg/m³
Drift	0,324	mg/m³
Pressure dependence	0,000	mg/m³
Ambient temperature dependence	0,650	mg/m³
Flow dependence	0,000	mg/m³
Voltage dependence	0,000	mg/m³
Uncertainty of test gas	0,381	mg/m³
Leakage during sampling and sample transport	0,000	mg/m³
Reference measuring method	0,426	mg/m³
Reproducibility standard deviation	0,375	mg/m³
Selectivity (cross interference):		
O2	0,000	mg/m³
CO	0,000	mg/m³
CO2	1,501	mg/m³
CH4	0,000	mg/m³
N2O	0,038	mg/m³
NO	0,000	mg/m³
NO2	0,000	mg/m³
NH3	0,000	mg/m³
SO2 (coal firing without desulfurization)	0,055	mg/m³
HCl (coal firing)	0,000	mg/m³
H2O (sample conditioning with cooler)	0,208	mg/m³

Result

Target value	< 6,6	mg/m³	according to 13. BlmSchV
Result 95% confidence interval	3,64	mg/m³	equivalent to s_{AMS} acc. to EN 14181
<i>equals the extended measurement uncertainty</i>			
Combined standard uncertainty	1,82	mg/m³	95% confidence interval met

Response Time

Target response time	< 200	s	
Measured response time	67	s	requirement fulfilled

Data base on: suitability test Ultramat 23 7MB2337, December 2004
Report-No: 427899, TÜV Industrie Service GmbH, TÜV Süd Gruppe

Manufacturer's Declaration of Conformity

for Automated Measuring Systems (AMS)

according to the requirements of EN 14956 and
QAL 1 according to EN 14181

SIEMENS AG A&D PI 2
76181 Karlsruhe, Germany


declares that the product

ULTRAMAT 23

7 MB 2338

CO 0-250 mg/m³

complies with the requirements of QAL 1 according to the international
standards EN 14956 and EN 14181 for the following specified
operating conditions:



Dr. Frank Diedrich
General Manager
A&D PI 2
Siemens AG
Datum: 15.06.2005




Peter Berghäuser
R&D Manager
A&D PI 2 RD
Siemens AG
Datum: 15.06.2005

Manufacturer's Declaration of Conformity for Automated Measuring Systems (AMS)
according to the requirements of EN 14956 and QAL 1 according to EN 14181

Specification of the Automated Measuring System

Gas analyzer
Order information
Measured component
Smallest TÜV certified measuring range



Range of Applications

Test gas concentration /	80	mg/m³
Emission limit value (daily average)	990 ... 1010	hPa
Ambient pressure range	20 ... 35	°C
Ambient temperature range	30 ... 90	l/h
Flow range	190 ... 250	V
Voltage range		

Determined Standard Uncertainties referred to Daily Average Limit Value

Non-linearity	0,289	mg/m³
Drift	0,600	mg/m³
Pressure dependence	0,000	mg/m³
Ambient temperature dependence	-1,269	mg/m³
Flow dependence	0,000	mg/m³
Voltage dependence	0,000	mg/m³
Uncertainty of test gas	0,924	mg/m³
Leakage during sampling and sample transport	0,000	mg/m³
Reference measuring method	0,516	mg/m³
Reproducibility standard deviation	0,922	mg/m³
Selectivity (cross interference):		
O2	0,000	mg/m³
CO	0,000	mg/m³
CO2	3,031	mg/m³
CH4	0,142	mg/m³
N2O	0,000	mg/m³
NO	0,000	mg/m³
NO2	0,124	mg/m³
NH3	0,000	mg/m³
SO2 (coal firing without desulfurization)	0,137	mg/m³
HCl (coal firing)	0,000	mg/m³
H2O (sample conditioning with cooler)	0,000	mg/m³

Result

Target value	< 8	mg/m³	according to 13. BlmSchV
Result 95% confidence interval	7,28	mg/m³	equivalent to s_{AMS} acc. to EN 14181
equals the extended measurement uncertainty			
Combined standard uncertainty	3,64	mg/m³	95% confidence interval net

Response Time

Target response time	< 200	s	
Measured response time	67	s	requirement fulfilled

Data base on: suitability test Ultramat 23 7MB233, August 1997
Report-No: 24012833, TÜV Umwelttechnik GmbH, TÜV Süddeutschland AG

Manufacturer's Declaration of Conformity

for Automated Measuring Systems (AMS)

according to the requirements of EN 14956 and
QAL 1 according to EN 14181

SIEMENS AG A&D PI 2
76181 Karlsruhe, Germany


declares that the product

ULTRAMAT 23

7 MB 2338

NO 0-400 mg/m³

complies with the requirements of QAL 1 according to the international
standards EN 14956 and EN 14181 for the following specified
operating conditions:



Dr. Frank Diedrich
General Manager
A&D PI 2
Siemens AG
Datum: 15.06.2005




Peter Berghäuser
R&D Manager
A&D PI 2 RD
Siemens AG
Datum: 15.06.2005

Manufacturer's Declaration of Conformity for Automated Measuring Systems (AMS)
according to the requirements of EN 14956 and QAL 1 according to EN 14181

Specification of the Automated Measuring System

Gas analyzer
Order information
Measured component
Smallest TÜV certified measuring range



Range of Applications

Test gas concentration /	131	mg/m ³
Emission limit value (daily average)	990 ... 1010	hPa
Ambient pressure range	990 ... 1010	hPa
Ambient temperature range	20 ... 35	°C
Flow range	30 ... 90	l/h
Voltage range	190 ... 250	V

Determined Standard Uncertainties referred to Daily Average Limit Value

Non-linearity	-0,462	mg/m ³
Drift	1,134	mg/m ³
Pressure dependence	0,000	mg/m ³
Ambient temperature dependence	4,681	mg/m ³
Flow dependence	0,000	mg/m ³
Voltage dependence	0,000	mg/m ³
Uncertainty of test gas	1,513	mg/m ³
Leakage during sampling and sample transport	0,000	mg/m ³
Reference measuring method	1,691	mg/m ³
Reproducibility standard deviation	1,325	mg/m ³
Selectivity (cross interference):		
O ₂	0,000	mg/m ³
CO	0,231	mg/m ³
CO ₂	3,695	mg/m ³
CH ₄	0,000	mg/m ³
N ₂ O	0,000	mg/m ³
NO	0,000	mg/m ³
NO ₂	0,000	mg/m ³
NH ₃	0,000	mg/m ³
SO ₂ (coal firing without desulfurization)	0,880	mg/m ³
HCl (coal firing)	0,000	mg/m ³
H ₂ O (sample conditioning with cooler)	0,000	mg/m ³

Result

Target value	< 26,2	mg/m ³	according to 13. BlmSchV
Result 95% confidence interval	13,39	mg/m ³	equivalent to s _{AMS} acc. to EN 14181
<i>equals the extended measurement uncertainty</i>			
Combined standard uncertainty	6,69	mg/m ³	95% confidence interval met

Response Time

Target response time	< 200	s	Requirement fulfilled
Measured response time	67	s	

Data base on: suitability test Ultramat 23 7MB2333, August 1997
Report-No: 24012833, TÜV Umwelttechnik GmbH, TÜV Süddeutschland AG

SIEMENS

Automation and Drives
Process Instrumentation and Analytics

Manufacturer's Declaration of Conformity

for Automated Measuring Systems (AMS)

according to the requirements of EN 14956 and
QAL 1 according to EN 14181

SIEMENS AG A&D PI 2
76181 Karlsruhe, Germany


declares that the product

ULTRAMAT 23

7 MB 2335, 7 MB 2337, 7 MB 2338

SO₂ 0-400 mg/m³

complies with the requirements of QAL 1 according to the international
standards EN 14956 and EN 14181 for the specified
operating conditions.



Dr. Frank Diedrich
General Manager
A&D PI 2
Siemens AG
Datum: 15.06.2005



Peter Berghäuser
R&D Manager
A&D PI 2 RD
Siemens AG
Datum: 15.06.2005

Manufacturer's Declaration of Conformity for Automated Measuring Systems (AMS)
according to the requirements of EN 14956 and QAL 1 according to EN 14181**Specification of the Automated Measuring System**

Gas analyzer
Order information
Measured component
Smallest TÜV certified measuring range

ULTRAMAT 23
7MB.2333 / 7MB.2337 / 7MB.2338
SO₂
0-100 mg/m³

Range of Applications

Test gas concentration /	130	mg/m ³
Emission limit value (daily average)	990 ... 1010	hPa
Ambient pressure range	990 ... 1010	hPa
Ambient temperature range	20 ... 35	°C
Flow range	30 ... 90	l/h
Voltage range	190 ... 250	V

Determined Standard Uncertainties referred to Daily Average Limit Value

Non-linearity	-0,462	mg/m ³
Drift	1,651	mg/m ³
Pressure dependence	0,000	mg/m ³
Ambient temperature dependence	-4,393	mg/m ³
Flow dependence	0,000	mg/m ³
Voltage dependence	0,000	mg/m ³
Uncertainty of test gas	1,501	mg/m ³
Leakage during sampling and sample transport	0,000	mg/m ³
Reference measuring method	1,678	mg/m ³
Reproducibility standard deviation	1,150	mg/m ³
Selectivity (cross interference):		
O ₂	0,000	mg/m ³
CO	0,000	mg/m ³
CO ₂	0,924	mg/m ³
CH ₄	0,906	mg/m ³
N ₂ O	0,616	mg/m ³
NO	0,000	mg/m ³
NO ₂	0,000	mg/m ³
NH ₃	0,462	mg/m ³
SO ₂ (coal firing without desulfurization)	0,000	mg/m ³
HCl (coal firing)	0,000	mg/m ³
H ₂ O (sample conditioning with cooler)	0,235	mg/m ³

Result

Target value	< 26	mg/m ³	according to 13. BlmSchV
Result 95% confidence intervall	11,13	mg/m ³	equivalent to s _{AMS} acc. to EN 14181
equals the extended measurement uncertainty			
Combined standard uncertainty	5,56	mg/m ³	95% confidence interval me

Response Time

Target response time	< 200	s	requirement fulfilled
Measured response time	67	s	

Data base on: suitability test Ultramat 23 7MB2333, August 1997
Report-No: 24012833, TÜV Umwelttechnik GmbH, TÜV Süddeutschland AG

TÜV Süddeutschland
TÜV Ecoplan Umwelt GmbH
Multi Component Process Gas Analyzers

ULTRAMAT 23

Report-Nr. 24012833

Manufacturer:

SIEMENS AG, Karlsruhe, Germany

TÜV Ecoplan Umwelt GmbH, TÜV Süddeutschland group, is herewith certifying for the Process gas analyzer ULTRAMAT 23, that the following measuring results for emission measurement at large power plants have been reached and are in accordance to the 13. ElmSchV, and TA-Luft regulations.

Smallest tested measuring ranges:

CO	0 ... 150 mg/m ³	NO	0 ... 250 mg/m ³
SO ₂	0 ... 400 mg/m ³	O ₂	0 ... 10 / 25 Vol.-%

Availability:

> 98% over a period of 3 months for two independant systems including sample preparation.

Drift:

With internal automatical Calibration with Ambient air:

Zero-drift : < 2% of range per year

Span-drift : < 2% of range per year

The performance of the internal autocalibration must be checked once a year.

Cross Interference:

The sum of all interferences to the measuring component stated above for SO₂, NO, O₂, NH₃, NO₂, CH₄, N₂O, CO, and H₂O with typical stack gas concentrations is < 4% of the ranges.

Limit of Detection:

CO	≤ 0,7% of range	NO	≤ 0,6% of range
SO ₂	≤ 0,8% of range	O ₂	≤ 0,04 Vol.-% O ₂

TÜV Ecoplan Umwelt GmbH
Unternehmensgruppe TÜV Süddeutschland
Westendstraße 199 D-80686 München
Tel. (089) 57 91 - 28 21, Fax (089) 57 91 - 28 22

Dr. A. Brandl *Dr. A. Witthuhn*
Dr. A. Brandl, München, September 1999 Dr. A. Witthuhn,
TÜV Ecoplan Umwelt GmbH, Westendstrasse 199, D-80686 München



Industrie Service

Certificate

TÜV Industrie Service GmbH
TÜV SÜD Gruppe

Laboratory for Environmental Services
(Laboratorium Umwelt Service)
accredited according DIN EN ISO/IEC 17025 DAP-PL-2885.99

Ultramat 23 - 7MB233
Gas Analyser for CO, NO, SO₂
Report Nr. 24012833 (August 1997)

Manufacturer:
Siemens AG, Karlsruhe, Germany

TÜV Industrie Service GmbH, TÜV SÜD Group is herewith certifying that the analyser Ultramat 23 7MB233 is in accordance with DIN EN ISO 14956, Jan. 2003 and fulfils QAL1 of EN 14181 for the following ranges of measurement or for higher ranges:

Component	C _{test} mg/ m ³	Range of measurement mg/ m ³	In accordance with DIN EN ISO 14956
CO	80	0-250	Yes
NO	80	0-250	Yes
NO as NO ₂	122	0-383	Yes
SO ₂	130	0-400	Yes

The calculation according DIN EN ISO 14956 was performed on the basis of the results of the investigations for report Nr. 24012833 (August 1997) for the German suitability test.

The following performance characteristics were regarded: Response time; Non-linearity, Instability/ drift; Selectivity/ interfering components; Dependence of ambient temperature, ambient pressure and voltage; Gas flow; Sample losses; Uncertainty of calibration gas; Uncertainty of independent reference method.

Munich, July 2004

H. Fiederer

Dr. D. Fiederer


Dr. A. Brandl

Laboratorium Umwelt Service, TÜV Industrie Service GmbH, Unternehmensgruppe TÜV Süd, IS-US1-MUC, Westendstrasse 199, D-80686 München

VERIFICATION DE REGLAGE / CALIBRATION TEST			Date:	08/01/2013
FDF101R6v5	Dernière rév :	27/04/2011	Lien procédure :	PROC010

Instruction(s) mise(s) en œuvre : 301 et 302

Analyseur d'oxygène / Oxygen analyser

N° de série capteur / SN Sensor	ATK/FD1301793
Longueur de pénétration / Length of penetration	400 mm
N° de série électronique / SN Control unit	ATK2020P13011821
Longueur du câble spécial / special cable length	40,00 m

Vérification de réglage / calibration test	Valeur nominale <i>Nominal value</i>	Valeur certificat <i>Certified value</i>	Fourchette d'acceptabilité <i>Precision</i>	N° Bouteille	N° Manomètre	Valeur affichée <i>Value on display</i>	Contrôle <i>Control</i>
20 % O2	20,90 %	+/- 0,30 %				21,00 %	PS
10 % O2	5,00 %	+/- 0,30 %		BV13780F	BT102	5,01 %	PS
0 % O2	0,40 %	+/- 0,30 %		BD62739F	BT102	0,41 %	PS
15 % O2	NA	+/- 0,30 %				NA	PS
5 % O2	5,00 %	+/- 0,30 %		BV13780F	BT102	5,01 %	PS
20 % O2	20,90 %	+/- 0,30 %				21,00 %	PS
10 % O2	5,00 %	+/- 0,30 %		BV13780F	BT102	5,02 %	PS
0 % O2	0,40 %	+/- 0,30 %		BD62739F	BT102	0,41 %	PS
15 % O2	NA	+/- 0,30 %				NA	PS
5 % O2	5,00 %	+/- 0,30 %		BV13780F	BT102	5,01 %	PS
20 % O2	20,90 %	+/- 0,30 %				20,90 %	PS

Options	
Alimentation / Power supply :	230V AC
Kit de râtroussillage / Back-flushing kit :	NA
Calorifuge / Heat insulator	OUI-YES
Contre-bride / Counter	Plate - Flat
Embout d'étalonnage / Calibration gas fitting device	OUI-YES
Ejecteur / Ejector	NON-NO
RS 232 / RS 232 port	NON-NO
Défaut de pression gaz test/Default of pressure of cal. Gas	NON-NO
Entrée 4-20mA mesure CO / CO 4-20mA Input	NON-NO



Instruments pour l'Environnement

Contrôlé par ACIME (Association pour la certification des instruments de Mesure pour l'Environnement), 1 rue Gaston Boissier, 75724 PARIS CEDEX 14

Référentiel : Marque NF379

N° Identification : 04/01/01-01

<i>Visa / Approved by</i>	
Cyril Capeau Resp. Production / Prod. Manager	

N° identification SETNAG / Identification No :	00379/12690
Date de vente / Date of sale :	09-janv-13
Client / Customer :	MICHELL INSTRUMENTS Ltd
Adresse / Address :	48 Lancaster Way Business Park ELY CAMBRIDGESHIRE CB6 3NW ROYAUME UNI



Cet article ne doit pas être mis au rebut avec les déchets municipaux normaux non-traités.
Mémo de nous contacter au 04.91.95.65.12 lorsqu'il arrive en fin de vie, nous nous occuperons de son recyclage.

Do not discard



Via Custoza, 31 - Chieti - www.laserlab.it - mail@laserlab.it
Tel. 0871 564343 - Fax. 0871 564443



TERMICA COLLEFERRO SPA

**Via Ariana km 5,2
00034 Colleferro (RM)**

ALLEGATO 7

Allegato 7 - Schema P&I laboratorio mobile, Certificati SRM TÜV/QAL1



CERTIFICATE

**TÜV Rheinland Immissionsschutz
und Energiesysteme GmbH**

Manufacturer: Horiba Europe GmbH


Measuring System: PG 250 SRM

Components: CO, NO_x, O₂

Test Report: 936/21206693/A, 2008-03-06

The measurement system fulfils
the requirements of
QAL 1
according to EN 14181 and EN ISO 14956.

Köln, 2008-06-20


Dr. rer. nat. Peter Wilbring


Dipl.-Chem. Martin Kerpa

www.umwelt-tuv.de / www.eco-tuv.com
tie@umwelt-tuv.de
Tel. +49 - 221 - 806 - 2275

TÜV Rheinland Immissionsschutz und Energiesysteme GmbH
Am Grauen Stein
51105 Köln

The company is accredited to DIN EN ISO/IEC 17025.

attached: 6 page(s)



EN ISO 14956 and prEN 15267-3 calculation for QAL 1 in EN 14181

Manufacturer data

Manufacturer

Horiba Europa GmbH

Measurement System

Emission measurement

Name

PG 250 SRM 1

Serial Number

H 0002ZBD

Measuring Principle

NDIR

TÜV Data

Approval Report

936/21206693/A

Date

06.03.2008

Editor

Guido Baum

Measurement Component and tested range

CO 75 mg/m³

Evaluation of the cross sensitivity (CS)

to 3 Vol.-% Oxygen
to 21 Vol.-% Oxygen
to 1 Vol.-% Humidity
to 300 mg/m³ Carbon monoxide
to 15 Vol.-% Carbon dioxide
to 57 mg/m³ Methane
to 40 mg/m³ Dinitrogen monoxide
to 100 mg/m³ Dinitrogen monoxide
to 300 mg/m³ Nitrogen monoxide
to 30 mg/m³ Nitrogen dioxide
to 20 mg/m³ Ammonia
to 200 mg/m³ Sulphur dioxide
to 1000 mg/m³ Sulphur dioxide
to 50 mg/m³ Hydrogen chloride
to 200 mg/m³ Hydrogen chloride

CS $X_{\text{max},j}$
0,00 mg/m³
0,00 mg/m³
0,00 mg/m³
0,00 mg/m³
0,00 mg/m³
0,61 mg/m³
0,00 mg/m³

0,01 mg/m³
0,00 mg/m³

Sum of positive cross sensitivities
Sum of negative cross sensitivities

Calculation of the combined standard uncertainty

Test Value

Lack of fit

$u(\Delta X_{\text{max},j}) = \frac{\Delta X}{\sqrt{3}}$

$u(\Delta X_{\text{max},j})^2$

Biggest interference (positiv or negativ)

0,152

Span shift in the field test

0,123

Zero shift in the field test

1,688

Sensitivity to sample volume flow

0,750

Sensitivity to ambient temperature

0,000

Dependence on supply voltage

0,610

Repeatability at span

0,030

Field reproducibility

0,001

Uncertainty of the test gas at the reference point

0,398

Combined standard uncertainty (u_c)

0,760

Total expanded uncertainty

2,098

Relative total expanded uncertainty

4,112

Requirement

8,2

Uc in % of the limit 50 mg/m³

10,0

Uc in % of the limit 50 mg/m³

Result: Requirements keep to QAL 1 of EN 14181



EN ISO 14956 and prEN 15267-3 calculation for QAL 1 in EN 14181

Manufacturer data

Manufacturer

Horiba Europe GmbH

Measurement System

Emission measurement

Name

PG 250 SRM 2

Serial Number

G 0800X2D

Measuring Principle

NDIR

TÜV Data

Approval Report

036/21206693/A

Date

06.03.2008

Editor

Guido Baum

Measurement Component and tested range

CO 76 mg/m³

Evaluation of the cross sensitivity (CS)

to 3 Vol.-% Oxygen
to 21 Vol.-% Oxygen
to 1 Vol.-% Humidity
to 300 mg/m³ Carbon monoxide
to 15 Vol.-% Carbon dioxide
to 57 mg/m³ Methane
to 40 mg/m³ Dinitrogen monoxide
to 100 mg/m³ Dinitrogen monoxide
to 300 mg/m³ Nitrogen monoxide
to 30 mg/m³ Nitrogen dioxide
to 20 mg/m³ Ammonia
to 200 mg/m³ Sulphur dioxide
to 1000 mg/m³ Sulphur dioxide
to 50 mg/m³ Hydrogen chloride
to 200 mg/m³ Hydrogen chloride

CS X_{max,J}
0,00 mg/m³
0,00 mg/m³
0,00 mg/m³
0,00 mg/m³
0,00 mg/m³
0,60 mg/m³
0,00 mg/m³

Sum of positive cross sensitivities
Sum of negative cross sensitivities

0,60 mg/m³
0,00 mg/m³

Calculation of the combined standard uncertainty

Test Value

$$u(\Delta X_{max,J}) = \frac{\Delta X}{\sqrt{3}} \quad n(\Delta X_{max,J})^2$$

Lack of fit

0,60 mg/m³ 0,39 mg/m³ 0,152

Biggest interference (positive or negative)

0,60 mg/m³ 0,35 mg/m³ 0,120

Span shift in the field test

2,25 mg/m³ 1,30 mg/m³ 1,688

Zero shift in the field test

2,10 mg/m³ 1,21 mg/m³ 1,470

Sensitivity to sample volume flow

0,00 mg/m³ 0,00 mg/m³ 0,000

Sensitivity to ambient temperature

1,89 mg/m³ 1,08 mg/m³ 1,191

Dependence on supply voltage

0,26 mg/m³ 0,14 mg/m³ 0,020

Repeatability at span

0,06 mg/m³ 0,03 mg/m³ 0,001

Field reproducibility

1,09 mg/m³ 0,63 mg/m³ 0,398

Uncertainty of the test gas at the reference point

1,50 mg/m³ 0,87 mg/m³ 0,750

Combined standard uncertainty (U_c)

$$U_c = \sqrt{\sum(u_{max,J})^2} \quad 2,406$$

Total expanded uncertainty

$$U_c = U_c * k \quad 4,716$$

Relative total expanded uncertainty

$$U_c \text{ in % of the limit } 50 \text{ mg/m}^3 \quad 9,4$$

Requirement

$$U_c \text{ in % of the limit } 50 \text{ mg/m}^3 \quad 10,0$$

Result: Requirements keep to QAL 1 of EN 14181



EN ISO 14956 and prEN 15267-3 calculation for QAL 1 in EN 14181

Manufacturer data

Manufacturer

Horiba Europe GmbH

Measurement System

Emulsion measurement

Name

PG 260 SRM 1

Serial Number

H 000228D

Measuring Principle

CLD

TÜV Data

Approval Report

936/21206693/A

Date

06.03.2008

Editor

Guido Baum

Measurement Component and tested range

NO 134 mg/m³

Evaluation of the cross sensitivity (CS)

to 3 Vol.-% Oxygen	CS X _{max,J}
to 21 Vol.-% Oxygen	0,00 mg/m ³
to 20 Vol.-% Humidity	-2,28 mg/m ³
to 300 mg/m ³ Carbon monoxide	0,00 mg/m ³
to 15 Vol.-% Carbon dioxide	0,00 mg/m ³
to 50 mg/m ³ Methane	0,67 mg/m ³
to 20 mg/m ³ Dinitrogen monoxide	0,00 mg/m ³
to 100 mg/m ³ Dinitrogen monoxide	0,00 mg/m ³
to 300 mg/m ³ Nitrogen monoxide	0,00 mg/m ³
to 30 mg/m ³ Nitrogen dioxide	0,00 mg/m ³
to 20 mg/m ³ Ammonia	0,00 mg/m ³
to 200 mg/m ³ Sulphur dioxide	0,00 mg/m ³
to 1000 mg/m ³ Sulphur dioxide	0,00 mg/m ³
to 50 mg/m ³ Hydrogen chloride	0,00 mg/m ³
to 200 mg/m ³ Hydrogen chloride	0,80 mg/m ³

Sum of positive cross sensitivities

1,47 mg/m³

Sum of negative cross sensitivities

-2,28 mg/m³

Calculation of the combined standard uncertainty

Test Value

Lack of fit

$$\Delta X_{\text{test},J} = \frac{\Delta X}{\sqrt{3}}$$

u(ΔX_{test,J})²

Biggest interference (positive or negative)

$$U_R = 1,47 \text{ mg/m}^3$$

0,724

Span shift in the field test

$$U_{f,1} = -2,28 \text{ mg/m}^3$$

1,730

Zero shift in the field test

$$U_{f,dr} = 1,74 \text{ mg/m}^3$$

1,012

Sensitivity to sample volume flow

$$U_{s,Vf} = 0,80 \text{ mg/m}^3$$

0,215

Sensitivity to ambient temperature

$$U_{s,T} = 0,00 \text{ mg/m}^3$$

0,000

Dependence on supply voltage

$$U_{s,U} = 6,39 \text{ mg/m}^3$$

13,618

Repeatability at span

$$U_{s,V} = 0,10 \text{ mg/m}^3$$

0,003

Field reproducibility

$$U_{s,D} = 0,12 \text{ mg/m}^3$$

0,005

Uncertainty of the test gas at the reference point

$$U_{s,ref} = 1,01 \text{ mg/m}^3$$

0,337

NOx converter efficiency adjustment

$$U_{s,NOx} = 2,68 \text{ mg/m}^3$$

2,394

Combined standard uncertainty (u_c)

$$U_c = 3,08 \text{ mg/m}^3$$

3,166

Total expanded uncertainty

$$u_c = \sqrt{\sum(u_{\text{max},j})^2}$$

4,817

Relative total expanded uncertainty

$$U_c = u_c * 1,98$$

9,442

Requirement

$$U_c \text{ in \% of the limit } 130,4 \text{ mg/m}^3$$

7,2

$$U_c \text{ in \% of the limit } 130,4 \text{ mg/m}^3$$

20,0

Result: Requirements keep to QAL 1 of EN 14181



EN ISO 14956 and prEN 15267-3 calculation for QAL 1 in EN 14181

Manufacturer data

Manufacturer

Measurement System

Name

Serial Number

Measuring Principle

Horiba Europe GmbH

Emission measurement

PG 250 SRM 2

G 0800X2D

CLD

TÜV Data

Approval Report

936/21206693/A

Date

08.03.2008

Editor

Guido Baum

Measurement Component and tested range

NO 134 mg/m³

Evaluation of the cross sensitivity (CS)

to 3 Vol.-% Oxygen
to 21 Vol.-% Oxygen
to 20 Vol.-% Humidity
to 300 mg/m³ Carbon monoxide
to 15 Vol.-% Carbon dioxide
to 50 mg/m³ Methane
to 20 mg/m³ Dinitrogen monoxide
to 100 mg/m³ Dinitrogen monoxide
to 300 mg/m³ Nitrogen monoxide
to 30 mg/m³ Nitrogen dioxide
to 20 mg/m³ Ammonia
to 200 mg/m³ Sulphur dioxide
to 1000 mg/m³ Sulphur dioxide
to 50 mg/m³ Hydrogen chloride
to 200 mg/m³ Hydrogen chloride

CS X_{max,J}
0,00 mg/m³
-2,55 mg/m³
0,00 mg/m³
0,00 mg/m³
0,67 mg/m³
0,00 mg/m³
1,07 mg/m³

Sum of positive cross sensitivities
Sum of negative cross sensitivities

1,74 mg/m³
-2,55 mg/m³

Calculation of the combined standard uncertainty

Test Value

Lack of fit

Biggest Interference (positiv or negativ)

Span shift in the field test

Zero shift in the field test

Sensitivity to sample volume flow

Sensitivity to ambient temperature

Dependence on supply voltage

Repeatability at span

Field reproducibility

Uncertainty of the test gas at the reference point

NOx converter efficiency adjustment

	$\Delta X_{max,J}$	$u(\Delta X_{max,J}) = \frac{\Delta X}{\sqrt{3}}$	$u(\Delta X_{max,J})^2$
U _{rl}	1,08 mg/m ³	1,08 mg/m ³	1,173
U _{nl}	-2,55 mg/m ³	-1,47 mg/m ³	2,161
U _{o,dr}	4,02 mg/m ³	2,32 mg/m ³	5,387
U _{a,dr}	1,07 mg/m ³	0,62 mg/m ³	0,383
U _{oM}	0,00 mg/m ³	0,00 mg/m ³	0,000
U _{ionp}	-6,50 mg/m ³	-3,75 mg/m ³	14,070
U _{ev}	0,13 mg/m ³	0,08 mg/m ³	0,006
U _{rep}	0,12 mg/m ³	0,07 mg/m ³	0,005
U _D	1,01 mg/m ³	0,68 mg/m ³	0,337
U _{adl}	2,68 mg/m ³	1,55 mg/m ³	2,394
U _{NOx}	3,48 mg/m ³	2,01 mg/m ³	4,046

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum(u_{max,J})^2} \quad 5,475$$

Total expanded uncertainty

10,730

Relative total expanded uncertainty

8,2

Requirement

20,0

Result: Requirements keep to QAL 1 of EN 14181



EN ISO 14956 and prEN 15267-3 calculation for QAL 1 in EN 14181

Manufacturer data

Manufacturer

Measurement System

Name

Serial Number

Measuring Principle

Horiba Europe GmbH

Emission measurement

PG 250 SRM 1

H 000228D

Paramagnetic

TÜV Data

Approval Report

936/21206693/A

Date

06.03.2008

Editor

Guldo Baum

Measurement Component and tested range

O₂ 25 Vol.-%

Evaluation of the cross sensitivity (CS)

	CS	X _{max,J}
to 3 Vol.-% Oxygen	0,00	Vol.-%
to 21 Vol.-% Oxygen	0,00	Vol.-%
to 30 Vol.-% Humidity	0,00	Vol.-%
to 300 mg/m ³ Carbon monoxide	0,00	Vol.-%
to 15 Vol.-% Carbon dioxide	-0,06	Vol.-%
to 50 mg/m ³ Methane	0,00	Vol.-%
to 20 mg/m ³ Dinitrogen monoxide	0,00	Vol.-%
to 100 mg/m ³ Dinitrogen monoxide	-0,03	Vol.-%
to 300 mg/m ³ Nitrogen monoxide	0,16	Vol.-%
to 30 mg/m ³ Nitrogen dioxide	0,00	Vol.-%
to 20 mg/m ³ Ammonia	0,00	Vol.-%
to 200 mg/m ³ Sulphur dioxide	0,00	Vol.-%
to 1000 mg/m ³ Sulphur dioxide	0,00	Vol.-%
to 50 mg/m ³ Hydrogen chloride	0,00	Vol.-%
to 200 mg/m ³ Hydrogen chloride	0,00	Vol.-%

Sum of positive cross sensitivities

0,16 Vol.-%

Sum of negative cross sensitivities

-0,09 Vol.-%

Calculation of the combined standard uncertainty

Test Value

Lack of fit

$$u(\Delta X_{\text{test},J}) = \frac{\Delta X}{\sqrt{3}}$$

$$u(\Delta X_{\text{max}})^2$$

Biggest interference (positiv or negativ)

$$0,003$$

Span shift in the field test

$$0,009$$

Zero shift in the field test

$$0,030$$

Sensitivity to sample volume flow

$$0,013$$

Sensitivity to ambient temperature

$$0,018$$

Dependence on supply voltage

$$0,018$$

Repeatability at span

$$0,000$$

Field reproducibility

$$0,006$$

Uncertainty of the test gas at the reference point

$$0,083$$

Combined standard uncertainty (u_c)

$$0,439$$

Total expanded uncertainty

$$0,861$$

Relative total expanded uncertainty

$$3,4$$

Requirement

$$6,0$$

$$\begin{aligned} u_c &= \sqrt{\sum(u_{\text{max},j})^2} \\ (u_c * k) &= u_c * 1,96 \end{aligned}$$

$$u_c = \sqrt{\sum(u_{\text{max},j})^2}$$

$$u_c = u_c * 1,96$$

$$u_c \text{ in % of the limit } 25 \text{ Vol.-%}$$

$$u_c \text{ in % of the limit } 25 \text{ Vol.-%}$$

Result: Requirements keep to QAL 1 of EN 14181

Attention: For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.



EN ISO 14956 and prEN 15267-3 calculation for QAL 1 in EN 14181

Manufacturer data

Manufacturer
Measurement System
Name
Serial Number
Measuring Principle

Horiba Europe GmbH
Emission measurement
PG 260 SRM 2
G 0800X2D
Paramagnetic

TÜV Data

Approval Report
Date
Editor

936/21206693/A
06.03.2008
Guido Baum

Measurement Component and tested range

O₂ 25 Vol.-%

Evaluation of the cross sensitivity (CS)

	CS X _{max,J}
to 3 Vol.-% Oxygen	0,00 Vol.-%
to 21 Vol.-% Oxygen	0,00 Vol.-%
to 30 Vol.-% Humidity	0,00 Vol.-%
to 300 mg/m ³ Carbon monoxide	0,03 Vol.-%
to 15 Vol.-% Carbon dioxide	-0,03 Vol.-%
to 50 mg/m ³ Methane	0,00 Vol.-%
to 20 mg/m ³ Dinitrogen monoxide	0,00 Vol.-%
to 100 mg/m ³ Dinitrogen monoxide	0,00 Vol.-%
to 300 mg/m ³ Nitrogen monoxide	0,06 Vol.-%
to 30 mg/m ³ Nitrogen dioxide	0,00 Vol.-%
to 20 mg/m ³ Ammonia	0,03 Vol.-%
to 200 mg/m ³ Sulphur dioxide	0,03 Vol.-%
to 1000 mg/m ³ Sulphur dioxide	0,00 Vol.-%
to 50 mg/m ³ Hydrogen chloride	-0,05 Vol.-%
to 200 mg/m ³ Hydrogen chloride	-0,15 Vol.-%

Sum of positive cross sensitivities

0,15 Vol.-%

Sum of negative cross sensitivities

-0,08 Vol.-%

Calculation of the combined standard uncertainty

Test Value

Lack of fit

$$u_{\text{fit}} = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,003

Biggest interference (positiv or negativ)

$$u_{\text{int}} = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,008

Span shift in the field test

$$u_{\text{span}} = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,030

Zero shift in the field test

$$u_{\text{zero}} = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,003

Sensitivity to sample volume flow

$$u_{\text{volum}} = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,018

Sensitivity to ambient temperature

$$u_{\text{temp}} = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,003

Dependence on supply voltage

$$u_{\text{voltage}} = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,000

Repeatability at span

$$u_{\text{repeat}} = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,006

Field reproducibility

$$u_{\text{field}} = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,083

Uncertainty of the test gas at the reference point

$$u_{\text{ref}} = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,003

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum u_{\text{max,J}}^2} \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,407

Total expanded uncertainty

$$U_a = u_c * 1,96 \quad n(u_{\Delta X_{\text{max,J}}})^2$$

0,798

Relative total expanded uncertainty

$$U_c \text{ in \% of the limit } 25 \text{ Vol.-%}$$

3,2

Requirement

$$U_c \text{ in \% of the limit } 25 \text{ Vol.-%}$$

6,0

Result: Requirements keep to QAL 1 of EN 14181

Attention: For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.



LM INDUSTRIES SA - DIVISION INSTRUMENTATION
46, Chemin de l'Eau, Case Postale 254, CH-1210 Cheseaux-Geneve, Suisse
Tel. +41 22 878 67 24 Fax +41 22 878 67 20 e-mail info@lmindustries.com
www.lm-instrument.com

Laboratory of calibration accredited ISO 17025 by the Swiss Accreditation Service



Collaborationssicher Kalibrierdienst
Service suisse d'accréditation
Service suisse de certification
Suisse Calibration Service

Nom d'identification : SCS 118
N° d'accréditation : Aperçu 118

The Swiss Accreditation Service is one of the signatories of the EA Mutual Agreement
for the recognition of calibration certificates

SCS

CERTIFICATE OF CALIBRATION

N° 960

SCS

Customer: Corporate Name:
Address: Laser Lab srl
Via Custoza, 31
69100 Chieri (CN)
Italy

Date of order: 22/09/2011
Order N°: CV603612

Device under test: Brand: SONIMIX 7000 2L2R
Serial N°: 3069
Produced by: LM INDUSTRIES SA
Measuring instruction: SONIMIX 7000 2L2R 3/27
Inlet pressure: 3000 mbar rel

Date of Calibration: 22/09/2011

This certificate of calibration confirms the link with the national standards which materialize the physical units (SI).

The results, uncertainties with confidence level and the methods of measurement are given in the following pages and belong to the certificate.

Stamp and Date

LM INDUSTRIES SA
Ch. de l'Eau 46
1210 Cheseaux/Geneve

23/03/14

For measurements
Delphine Jaouen

Chief of the laboratory of calibration
J.W. Berthelot

This certificate of calibration should not be reproduced, if not completely, without the written authorization of the laboratory.
Certificate N°: 960

Page 1 sur 4

KRTECORA

COMPANY
WITH QUALITY MANAGEMENT
SYSTEM CERTIFIED BY DNV
• IEC 9001:2000

R.Tecora s.r.l. - 20094 Corsico - Milano - Via A. Volta, 22 - Tel +39 02 4505501 - Fax +39 0248601811 - www.tecora.it
R.Tecora s.r.l. - 20094 Corsico - Milano - Via A. Volta, 22 - Tel +39 02 4505501 - Fax +39 0248601811 - www.tecora.it
Reg. Imp. 191048 Trib di Milano - Cod. Fisca. R.P. N° IT04579990151
N.I.A.A. 1023629 di Milano - N. Rob. M. 1037198 - Reg. Imp. 191048 Trib di Milano - Cod. Fisca. R.P. N° IT04579990151

Strumento - Instrument: Icostatik Basic HV
Matricola - Serial Number: 0098397
Anno di costruzione - Construction Year: 2010

**Dichiarazione di conformità
Certificate of Conformity**

In qualità di costruttore dichiariamo sotto la nostra responsabilità che
l'apparecchiatura di cui sopra è conforme alle direttive 2006/95/CE, 2004/108/CE, 2006/42/CE,
DPR 459-96 ed è stata realizzata in conformità
alle norme armonizzate applicabili CEEI EN 61010-1

As manufacturer we declare under our sole responsibility that the equipment,
in accordance with the provisions of the Directives 2006/95/CE, 2004/108/CE, 2006/42/CE,
and with the harmonized norm EN 61010-1

Date: 09/03/2010



Via Custoza, 31 - Chieti - www.laserlab.it - mail@laserlab.it
Tel. 0871 561443 - Fax. 0871 561443



DRYCAL
ISO 9001
n° S00016

DRYCAL
ISO 14001
n° A00203

Laser Lab	PO-71/1
Procedura Operativa n°71	Rev. 01 del 23/10/2009
Titolo della procedura: Taratura interna miscelatore Sonimix 7000-2L2R	Pag. 1 di 4


CERTIFICATO DI TARATURA					
CLIENTE STRUMENTO TESTATO UNITÀ GAS S/N PROCEDURA DI RIFERIMENTO	LASER LAB				
	SONIMIX 7000-2L2R				
	3366				
	PO-71				
FLUSSO IN CONDIZIONI NORMALI MPC CAR 0,25...5,0 l/min (25±1°C, 1013 hPa) MISCELA GAS TECNICI COMPLEMENTARI AZOTO certificato analisi N° 13006 calibrato con DRYCAL DC-LITE DCL-MH SN.106723	IMPOSTATO	RILEVATO (Drycal)	MISURATO (Miscelatore)	INCERTEZZA TOT RELATIVA %	CRITERIO DI ACCETTABILITÀ
	ml/min				
	250,00	255,27	252,31	0,73	2
	500,00	552,66	526,36	1,55	2
	1000,00	1047,43	1018,45	0,15	2
	1500,00	1628,09	1540,23	0,73	2
	2000,00	2155,79	2031,65	0,66	2
	2500,00	2728,58	2523,11	0,96	2
	3000,00	3172,57	3089,54	0,39	2
	3500,00	3750,18	3545,22	0,82	2
FLUSSO IN CONDIZIONI NORMALI MPC DIL 1 0,25...5,0 l/min (25±1°C, 1013 hPa) MISCELA GAS TECNICI COMPLEMENTARI AZOTO certificato analisi N° 13006 calibrato con DRYCAL DC-LITE DCL-MH SN.106723	4000,00	4157,11	4033,81	0,73	2
	4500,00	5046,26	4722,02	0,25	2
	ml/min				
	250,00	277,07	252,14	0,62	2
	500,00	552,17	523,65	0,51	2
	1000,00	1205,72	1089,31	0,29	2
	1500,00	1624,11	1487,2	0,38	2
	2000,00	2247,14	2049,01	0,51	2
	2500,00	2681,02	2421,58	0,85	2
	3000,00	3149,70	2929,54	0,13	2
	3500,00	3711,31	3411,68	0,47	2
	4000,00	4223,39	4102,32	0,20	2
	5000,00	5123,13	4756,9	0,17	2

Laser Lab S.r.l.

Via Custoza, 31 Chieti (CH)
66100 ITALY

DATA

16/05/2013





Via Custoza, 31 - Chieti - www.laserlab.it - mail@laserlab.it
Tel. 0871 564343 - Fax. 0871 564443



TERMICA COLLEFERRO SPA
Via Ariana km 5,2
00034 Colleferro (RM)

Allegato 8

Elenco prove accreditate e Certificato di accreditamento Accredia

CERTIFICATO DI ACCREDITAMENTO *Accreditation Certificate*

Accreditamento n°
Accreditation n°

0142

Rev. 1

Si dichiara che
We declare that

LASER LAB Srl

Sede:
Via Custoza 31 - 66013 Chieti CH

è conforme ai requisiti
della norma

UNI CEI EN ISO/IEC 17025:2005 "Requisiti generali per la competenza dei
Laboratori di prova e taratura"

*meets the requirements
of the standard*

*EN ISO/IEC 17025:2005 "General Requirements for the Competence of Testing
and Calibration Laboratories" standard*

quale **Laboratorio di Prova**

as **Testing Laboratory**

L'accreditamento attesta la competenza tecnica del Laboratorio relativamente allo scopo riportato nelle schede indicate al presente certificato. Le schede possono variare nel tempo. I requisiti gestionali della ISO/IEC 17025:2005 (sezione 4) sono scritti in un linguaggio idoneo all'attività del Laboratorio di Prova, sono conformi ai principi della ISO 9001:2008 ed allineati con i suoi requisiti applicabili.

Il presente certificato non è da ritenersi valido se non accompagnato dalle schede indicate e può essere sospeso o revocato in qualsiasi momento nel caso di inadempienza accertata da parte di ACCREDIA. La vigenza dell'accreditamento può essere verificata sul sito WEB (www.accredia.it) o richiesta direttamente ai singoli Dipartimenti.

The accreditation certifies the technical competence of the laboratory limited to the scope detailed in the attached Enclosure. The scope may vary in the time. The management system requirements in ISO/IEC 17025:2005 (Section 4) are written in a language relevant to Testing Laboratories operations and meet the principles of ISO 9001:2008 and are aligned with its pertinent requirements.

The present certificate is valid only if associated to the annexed schedule, and can be suspended or withdrawn at any time in the event of non fulfilment as ascertained by ACCREDIA.

The in force status of the accreditation may be checked in the WEB site (www.accredia.it) or on direct request to appointed Department.

Data di 1ª emissione
1st issue date
1997-04-03

Data di modifica
Modification date
2013-07-16

Data di scadenza
Expiring date
2017-05-04

Il Direttore Generale
The General Director
(Dr. Filippo Trifiletti)

Il Direttore di Dipartimento
Department Director
(Dr. Paolo Bianco)

Il Presidente
The President
(Cav. del Lav. Federico Grazioli)

LASER LAB Srl Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A	
	Revisione: 25	Data: 16/07/2013
	Scheda 1 di 20	PA163AR25.pdf

ELENCO PROVE ACCREDITATE - CATEGORIA: 0

Acque naturali e di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Carbonio organico disciolto (DOC)	APAT CNR IRSA 5040 Man 29 2003
Oli e Grassi animali e vegetali (da calcolo)	APAT CNR IRSA 5160 B1 Man 29 2003 + APAT CNR IRSA 5160 B2 Man 29 2003

Acque naturali e di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Idrocarburi totali	APAT CNR IRSA 5160 B2 Man 29 2003

Acque naturali e di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Sostanze oleose totali	APAT CNR IRSA 5160 B1 Man 29 2003

Acque naturali, superficiali, sotterranee, di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Cloruri, Fluoruri, Nitrati, Solfati, Fosfati, Bromuri	APAT CNR IRSA 4020 Man 29 2003

Acque destinate al consumo umano

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Colore	Rapporti ISTISAN 2007/31 pag 90 Met ISS BJA 021
Odore	Rapporti ISTISAN 2007/31 pag 80 Met ISS BAA 026
Residuo fisso a 180°C	Rapporti ISTISAN 2007/31 pag 65 Met ISS BFA 032
Sapore	Rapporti ISTISAN 2007/31 pag 85 Met ISS BKA 028

Acque destinate al consumo umano, naturali e di piscina

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Azoto Organico, Azoto Kjeldahl	APAT CNR IRSA 5030 Man 29 2003
Cloro libero e Cloro totale	Rapporti ISTISAN 2007/31 pag 45 Met ISS BHD 033
Colore	APAT CNR IRSA 2020 A Man 29 2003
Conducibilità elettrica	Rapporti ISTISAN 2007/31 Pag 55 Met ISS BDA 022
pH	Rapporti ISTISAN 2007/31 Pag 68 Met ISS BCA 023

Acque destinate al consumo umano, naturali, minerali, di pozzo, di falda e di piscina

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Indice di permanganato (Ossidabilità Kubel)	UNI EN ISO 8467: 1997

Acque destinate al consumo umano, naturali, superficiali, sotterranee

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Conta di Pseudomonas aeruginosa	UNI EN ISO 16266: 2008
Numerazione di Clostridium perfringens	DLgs n° 31 02/02/2001 GU SO n° 52 03/03/2001 All III
Ricerca ed enumerazione di enterococchi intestinali.	UNI EN ISO 7899-2: 2003
Ricerca ed enumerazione di Escherichia coli e batteri coliformi	UNI EN ISO 9308-1: 2002

Acque destinate al consumo umano, naturali, superficiali, sotterranee

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
---	------------------------

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A	
	Revisione: 25	Data: 16/07/2013
	Scheda 2 di 20	PA163AR25.pdf

Numerazione di microrganismi coltivabili: conteggio delle colonie a 22°C e 37°C UNI EN ISO 6222: 2001

Acque destinate al consumo umano, naturali, superficiali, sotterranee, di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Carbonio Organico Totale (TOC), Carbonio organico dissolto (DOC)	UNI EN 1484: 1999

Acque di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Azoto ammoniacale	APAT CNR IRSA 4030 A2 Man 29 2003
Policlorobifenili (PCB): Aroclor 1260, Aroclor 1254, Aroclor 1242	APAT CNR IRSA 5110 Man 29 2003

Acque di scarico, superficiali e sotterranee

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Glicoli: Glicole etilenico, glicole dietilenico, glicole trietilenico	M.U. 1367:99
Mercurio	UNI EN 1483:2008
Odore	APAT CNR IRSA 2050 Man 29 2003
Tensioattivi anionici	APAT CNR IRSA 5170 Man 29 2003

Acque e campioni acquosi

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Idrocarburi policiclici aromatici (IPA): naftalene, acenaftilene, acenaftene, fluorene, fenantrene, antracene, fluorantene, pirene, benzo(a) antracene, crisene, benzo(b) fluorantene, benzo(a) pirene, dibenzo(a,h) antracene, benzo(g,h,i) perilene, indeno(1,2,3,cd) pirene, benzo(k) fluorantene	APAT CNR IRSA 5080 Man 29 2003 (ad esclusione del paragrafo 7.4 e 7.1.2)

Acque e Campioni acquosi

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Acrilammide	DIN 38413-6 2007
Aldeidi: formaldeide, acetaldeide, acroleina, propionaldeide, butirraldeide, benzaldeide, gluteraldeide	EPA 8315A 1996
Alluminio, Antimonio, Argento, Arsenico, Berillio, Boro, Cadmio, Calcio, Cobalto, Cromo, Ferro, Fosforo, Magnesio, Manganese, Mercurio, Molibdeno, Nichel, Piombo, Potassio, Rame, Selenio, Sodio, Stronzio, Tallio, Tellurio, Vanadio, Zinco.	EPA 3015A 2007+ EPA 6010C 2007
Alluminio, Antimonio, Argento, Arsenico, Berillio, Cadmio, Ferro, Manganese, Nichel, Piombo, Selenio, Tallio, Vanadio, Cobalto, Cromo, Rame, Zinco, Boro; Bario; Molibdeno.	EPA 3015A 2007+ EPA 6020A 2007
Alluminio, Antimonio, Argento, Arsenico, Berillio, Cadmio, Ferro, Manganese, Nichel, Piombo, Selenio, Tallio, Vanadio, Cobalto, Cromo, Rame, Zinco, Boro; Bario; Molibdeno.	EPA 6020A 2007
Alluminio, Antimonio, Argento, Bario, Berillio, Boro, Cadmio, Calcio, Cobalto, Cromo, Ferro, Fosforo, Magnesio, Manganese, Nichel, Piombo, Potassio, Rame, Selenio, Silice, Sodio, Stagno, Stronzio, Tallio, Vanadio, Zinco, Titanio, Litio, Arsenico, Mercurio, Molibdeno, Tellurio	EPA 6010C 2007
Alluminio, Arsenico, Bario, Berillio, Cadmio, Calcio, Cobalto, Cromo, Ferro, Magnesio, Manganese, Nichel, Piombo, Potassio, Rame, Selenio, Sodio, Tallio, Vanadio, Zinco, Molibdeno	EPA 3010A 1992 + EPA 6010C 2007
Ammine aromatiche: anilina, o-anisidina, m-anisidina, p-anisidina, difenilammina, p-toluidina	EPA 3510C 1996 + EPA 8270D 2007
Atrazina, Alaclor (> 0,01 µg/l)	MP 279/C rev 5 2013
Bromuri, Cloruri, Floruri, Nitrati, Fosfati, Solfati	EPA 9056A 2007

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A	
	Revisione: 25	Data: 16/07/2013
	Scheda 3 di 20	PA163AR25.pdf

Cianuri (liberi e totali)	ISO 6703-2:1984 sez. 1 e 2
Cloruri , Salinità (come NaCl)	APHA Standard Methods for the Examination of Water and Wastewater ed 22nd 2012 4500 B
Composti organici alogenati: cloroformio, tetrachloruro di carbonio, 1,1,1,1 tricloroetano, tricloroetilene, tetrachloroetilene, 1,1,2,2 tetrachloroetano, 1,1,1,2 tetrachloroetano, clorometano, 1,2 dicloroetano, esaclorobutadiene, 1,1 dicloroetilene, 1,1 dicloroetano, 1,2 dicloropropano, 1,1,2 tricloroetano, 1,2,3 tricloropropano, cis-1,2 dicloroetilene, trans-1,2 dicloroetilene, 1,2 dibromoetano monoclorobenzene, 1,2 diclorobenzene, 1,4 diclorobenzene, bromoformio, bromodicitrormetano, dibromoclorometano, esaclorobenzene, pentaclorobenzene, 1,2,4 triclorobenzene, 1,2,4,5 tetrachlorobenzene.	EPA 5021A 2003 + EPA 8021B 1996
Idrocarburi < C12 (come sommatoria C6-C12), Idrocarburi =< C12 (come sommatoria C6-C12)	EPA 5021A 2003 +EPA 8015C 2007
Idrocarburi > C12 (come sommatoria C13-C40)	EPA 3510C 1996 + EPA 8015C 2007
Idrocarburi policiclici aromatici (IPA): naftalene, acenaftilene, acenaftene, fluorene, fenantrene, antracene, fluorantene, pirene, benzo(a) antracene, crisene, benzo(b) fluorantene, benzo(a) pirene, dibenzo(a,h) antracene, benzo(g,h,i) perilene, indeno(1,2,3,cd) pirene, benzo(k) fluorantene; Nitrobenzeni: Nitrobenzene, 2-cloronitrobenzene, 3- cloronitrobenzene, 4-cloronitrobenzene 1,3-dinitrobenzene, 1,2-dinitrobenzene; Clorobenzeni: 1,2,4-triclorobenzene, 1,2,4,5-tetrachlorobenzene, pentaclorobenzene, esaclorobenzene, Ftalati: Bis(2-etilesil)ftalato, butil benzil ftalato, di-n-butil ftalato, di-n-ottile ftalato, dietil ftalato, dimetil ftalato; acido paraftalico Fenoli: Fenolo, m+p cresolo, o-cresolo, 2-clorofenolo, 2,4-diclorofenolo, 2,4,6-triclorofenolo, pentaclorofenolo	EPA 3510C 1996+ EPA 8270D 2007
Idrocarburi totali (C<12 nel range C5 -C12 + C>12 nel range C13-C40), Idrocarburi totali (espressi come n-esano)	EPA 5021A 2003 +EPA 8015C 2007+ EPA 3510C 1996 + EPA 8015C 2007
Metilterbutiletere (MTBE), Solventi organici aromatici: benzene, toluene, o-m-p-xilene, etilbenzene, stirene, cumene.	EPA 5021A 2003 + EPA 8015C 2007
Pesticidi organofosforati: Clorpirimifos-me, Clorpirimifos-et, Pirimifos-me , Paration-me, disulfoton. Pesticidi clorurati: esaclorobenzene, a-HCH, b-HCH, aldrina, 4,4'-Diclorodifeniltricloroetano (4,4'-DDT), 2,4'-Diclorodifeniltricloroetano (2,4'-DDT), 4,4'-Diclorodifenildicloroetano (4,4'-DDD), 2,4'-Diclorodifenildicloroetano (2,4'-DDD), 4,4'-Diclorodifenildicloroetilene (4,4'-DDE), 2,4'-Diclorodifenildicloroetilene (2,4'-DDE), somma di DDD, DDT e DDE, lindano (g-HCH), dieldrin, endrin, isodrin, clordano.	EPA 3510C 1996 + EPA 8270D 2007
Piombo tetraetile	EPA 3510C 1996 + EPA 8270D 2007
Policlorobifenili (PCB): #28, #52, #77, #81, #95, #99, #101, #105, #110, #114, #118, #123, #126, #128+#167, #138, #146, #149, #151, #153, #156, #157, #169, #170, #177, #180, #183, #187, #189; Polichlorobifenili (PCB) Diossina-simili: #77, #81, #105, #114, #118, #123, #126, #156, #157, #167, #169, #189; PCB Totali	EPA 3510C 1996 + EPA 8270D 2007
Potenziale Redox	APHA Standard Methods for the Examination of Water and Wastewater ed 22nd 2012 2580 B
Richiesta biochimica d'ossigeno (BOD5)	APHA Standard Methods for the Examination of Water and Wastewater ed 22nd 2012 5210 D
Richiesta chimica di ossigeno (COD)	ISO 15705: 2002
Solidi fissi e volatili a 600°C	APAT CNR IRSA 2090 D Man 29 2003
Solventi organici aromatici: benzene, toluene, m,p-xilene, o-xilene, etilbenzene, stirene, cumene	APAT CNR IRSA 5140 Man 29 2003

LASER LAB SrL	Numero di accreditamento: 0142 Sede A
Via Custoza 31 66013 Chieti CH	Revisione: 25 Data: 16/07/2013
	Scheda 4 di 20 PA163AR25.pdf

Solventi organici clorurati e alogenati: cloroformio, tetrachloruro di carbonio, 1,1,1-tricloroetano, tricloroetilene, tetrachloroetilene, 1,1,2,2-tetrachloroetano, 1,1,1,2-tetrachloroetano, clorometano, diclorometano, 1,2-dicloroetano, 1,1-dicloroetilene, 1,1-dicloroetano, 1,2-dicloropropano, 1,1,2-tricloroetano, 1,2,3-tricloropropano, cis-1,2 dicloroetilene, 1,2-dibromoetano, monoclorobenzene, 1,2-diclorobenzene, 1,4-diclorobenzene, bromoformio, bromodichlorometano, dibromoclorometano

APAT CNR IRSA 5150 Man 29 2003

Solventi organici clorurati e alogenati: cloroformio, tetrachloruro di carbonio, 1,1,1tricloroetano, tricloroetilene, tetrachloroetilene, 1,1,2,2tetrachloroetano, 1,1,1,2tetracloroetano, clorometano, diclorometano, cloruro di vinile monomero (CVM), 1,2dicloroetano, esaclorobutadiene, 1,1-dicloroetilene, 1,1-dicloroetano, 1,2dicloropropano, 1,1,2-tricloroetano, 1,2,3-tricloropropano, cis-1,2 dicloroetilene, trans-1,2 dicloroetilene, 1,2-dibromoetano, 1,3-diclorobenzene, monoclorobenzene, 1,2-diclorobenzene, 1,4-diclorobenzene, bromoformio, dibromodichlorometano, bromodichlorometano, diclorobromometano. Solventi organici aromatici: benzene, toluene, p-o-m-xilene, etilbenzene, stirene, cumene. Trialometani Totali (cloroformio, bromoformio, dibromoclorometano, bromodichlorometano)

EPA 5030C 2003 + EPA 8260C 2006

Acque e campioni acquosi, Sedimenti, Suoli

Denominazione della prova / Campi di prova

Dibenzodiossine/furani policlorurati (PCDD/PCDF):
2,3,7,8-Tetraclorodibenzodiossina (TCDD),
1,2,3,7,8-Pentaclorodibenzodiossina (PeCDD),
1,2,3,4,7,8-Esaclorodibenzodiossina (HxCDD),
1,2,3,6,7,8-Esaclorodibenzodiossina (HxCDD),
1,2,3,7,8,9-Esaclorodibenzodiossina (HxCDD),
1,2,3,4,6,7,8-Eptaclorodibenzodiossina (HpCDD), Octaclorodibenzodiossina (OCDD) Policlorodibenzofurani (PCDF): 2,3,7,8-Tetraclorodibenzofurano (TCDF), 1,2,3,7,8-Pentaclorodibenzofurano (PeCDF), 2,3,4,7,8-Pentaclorodibenzofurano (PeCDF),
1,2,3,4,7,8-Esaclorodibenzofurano (HxCDF),
1,2,3,6,7,8-Esaclorodibenzofurano (HxCDF),
1,2,3,7,8,9-Esaclorodibenzofurano (HxCDF),
2,3,4,6,7,8-Esaclorodibenzofurano (HxCDF),
1,2,3,4,6,7,8-Eptaclorodibenzofurano (HpCDF), 1,2,3,4,7,8,9-Eptaclorodibenzofurano (HpCDF), Octaclorodibenzofurano (OCDF) ;
Sommatoria PCDD/PCDF I-TEQ

Metodo di prova

EPA 1613 B 1994 + NATO/CCMS Report n° 176
1988

Acque naturali e di scarico

Denominazione della prova / Campi di prova

Acidità e Alcalinità (Idrossidi, Carbonati, Bicarbonati, Alcalinità totale)

Metodo di prova

APAT CNR IRSA 2010 B Man 29 2003

Aldeidi

APAT CNR IRSA 5010 A Man 29 2003

Cloro attivo libero, cloro residuo

APAT CNR IRSA 4080 Man 29 2003

Cromo esavalente (Cromo VI)

APAT CNR IRSA 3150 C Man 29 2003

Cromo esavalente (Cromo VI)

APAT CNR IRSA 3150 B2 Man 29 2003

Fenoli: fenolo, fenoli orto e meta-sostituiti e para-sostituiti con sostituente carbossile, alogeno, metossile o gruppo solfonico.

APAT CNR IRSA 5070 A1 Man 29 2003, APAT CNR IRSA 5070 A2 Man 29 2003

pH

APAT CNR IRSA 2060 Man 29 2003

Richiesta chimica d'ossigeno (COD)

APAT CNR IRSA 5130 Man 29 2003

Solfito

APAT CNR IRSA 4150 A Man 29 2003

Solidi Sedimentabili

APAT CNR IRSA 2090 C Man 29 2003

Solidi sospesi totali

APAT CNR IRSA 2090 B Man 29 2003

Solidi Totali disciolti (Residuo secco a 105 °C)

APAT CNR IRSA 2090 A Man 29 2003

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A	
	Revisione: 25	Data: 16/07/2013
	Scheda 5 di 20	PA163AR25.pdf

Acque naturali, destinate al consumo umano e industriali

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Durezza totale (come CaCO ₃)	APAT CNR IRSA 2040 B Man 29 2003

Acque naturali, di mare e di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Azoto nitroso	APAT CNR IRSA 4050 Man 29 2003

Acque naturali, dolci, di mare, sotterranee

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Azoto ammoniacale	APAT CNR IRSA 4030 A1 Man 29 2003

Acque potabili, superficiali e di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Torbidità	APAT CNR IRSA 2110 Man 29 2003

Acque reflue, superficiali, sotterranee

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Alluminio, Arsenico, Bario, Berillio, Cadmio, Calcio, Cobalto, Cromo, Ferro, Magnesio, Manganese, Nichel, Piombo, Potassio, Rame, Selenio, Sodio, Tallio, Tellurio, Vanadio, Zinco, Molibdeno.	APAT CNR IRSA 3020 Man 29 2003

Acque sotterranee, superficiali, di mare e di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Conducibilità elettrica	APAT CNR IRSA 2030 Man 29 2003

Acque sotterranee, superficiali, di mare, di scarico, domestici ed industriali

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Solfuro	APAT CNR IRSA 4160 Man 29 2003

Acque superficiali, di fiume, di lago ed acque di scarico anche sottoposte a trattamento

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Conta Spore di clostridi solfito riduttori	APAT CNR IRSA 7060 B Man 29 2003
Conta Streptococchi fecali, Enterococchi (MF)	APAT CNR IRSA 7040 C Man 29 2003
Conta Coliformi fecali (MF)	APAT CNR IRSA 7020 B Man 29 2003
Conta Coliformi fecali (MPN)	APAT CNR IRSA 7020 A Man 29 2003
Conta Coliformi totali (MF)	APAT CNR IRSA 7010 C Man 29 2003
Conta Coliformi totali (MPN)	APAT CNR IRSA 7010 A Man 29 2003
Conta Escherichia coli	APAT CNR IRSA 7030 C Man 29 2003
Conta Escherichia coli	APAT CNR IRSA 7030 F Man 29 2003
Conta Streptococchi fecali, Enterococchi (MPN)	APAT CNR IRSA 7040 A Man 29 2003
Conteggio delle colonie su Agar a 36 °C e 22 °C	APAT CNR IRSA 7050 Man 29 2003

Alimenti

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Ceneri	Rapporti ISTISAN 1996/34 Pag 77
Cloruro di sodio (>0,5 % (m/m))	MP 65/C rev 3 2013
Piombo, Cadmio	UNI EN 13805: 2002 + UNI EN 14083: 2003
Sostanze azotate totali, Proteine (da calcolo)	Rapporti ISTISAN 1996/34 pag 13

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A
	Revisione: 25 Data: 16/07/2013
	Scheda 6 di 20 PA163AR25.pdf

Sostanze grasse totali	Rapporti ISTISAN 1996/34 pag 41 Met A
Sostanze grasse totali	Rapporti ISTISAN 1996/34 pag 39
Alimenti che non contengono sostanze termolabili a 103°C	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Umidità (Residuo secco)	Rapporti ISTISAN 1996/34 Pag 7 Met B
Alimenti destinati al consumo umano ed animale	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Conta Bacillus Cereus presunto a 30°	UNI EN ISO 7932:2005
Conta Batteri lattici mesofili	NF ISO 15214:1998
Conta Batteri solfito riduttori	NF V 08-061 2009
Conta Escherichia coli beta-glucuronidasi positivo	ISO 16649-2:2001
Conta Listeria monocytogenes	UNI EN ISO 11290-2:2005
Conta Listeria monocytogenes	NF EN ISO 11290-2:1998/A1: 2005
Conta Stafilococchi coagulasi positivi (Staphylococcus aureus e altre specie)	UNI EN ISO 6888-1:2004
Conta Stafilococchi coagulasi positivi (Staphylococcus aureus e altre specie)	UNI EN ISO 6888-2:2004
Ricerca Listeria monocytogenes	NF EN ISO 11290-1:1997/A1: 2005
Alimenti destinati al consumo umano ed animale (con aw maggiore di 0,95)	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Conta Lieviti e Muffe	ISO 21527-1:2008
Alimenti destinati al consumo umano ed animale (con aw minore o uguale di 0,95)	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Conta Lieviti e Muffe	ISO 21527-2:2008
Alimenti destinati al consumo umano ed animale (escluso latte e derivati)	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Listeria monocytogenes	Rapporti ISTISAN 1996/35 Met 20
Alimenti destinati al consumo umano ed animale, tamponi	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Conta Coliformi	ISO 4832:2006
Conta Coliformi a 44°C	NF V 08-060 2009
Conta Coliformi a 30°C	NF V 08-050 2009
Conta Enterobacteriaceae	ISO 21528-2: 2004
Conta Lieviti e Muffe	NF V 08-059 2002
Conta microbica a 30°C	ISO 4833:2003
Conta Stafilococchi coagulasi positivi	NF V 08-057-1 2004
Ricerca Listeria monocytogenes	UNI EN ISO 11290-1:2005
Alimenti grassi: Oli e grassi	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A
	Revisione: 25 Data: 16/07/2013
	Scheda 7 di 20 PA163AR25.pdf

Multiresiduale dei fitofarmaci clorurati: alfa-HCH, beta-HCH, Aldrin, 4,4'-Diclorodifeniltricloroetano (4,4'-DDT), 2,4'-Diclorodifeniltricloroetano (2,4'-DDT), 4,4'-Diclorodifenildicloroetano (4,4'-DDD), 2,4'-Diclorodifenildicloroetano (2,4'-DDD), 4,4'-Diclorodifenildicloroetilene (4,4'-DDE), 2,4'-Diclorodifenildicloroetilene (2,4'-DDE), somma di DDD, DDT e DDE, Lindano (gamma-HCH), Dieldrin, Endrin, Eptacloro, Eptacloroepossido, Clordano

UNI EN 1528-1: 1997 + UNI EN 1528-2: 1997 + UNI EN 1528-3: 1997 Met F + UNI EN 1528-4: 1997

Aria: Ambienti di lavoro

Denominazione della prova / Campi di prova	Metodo di prova
Cromo esavalente (Cromo VI)	NIOSH 7600 1994
Fibre di Amianto aerodisperse	DM 06/09/1994 GU SO n° 288 10/12/1994 AlI 2A

Aria: Ambienti di lavoro, Fiala attivata (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Aldeidi: formaldeide, acetaldeide, propionaldeide, butirraldeide, benzaldeide, acroleina	EPA 0100 1996 + EPA 8315A 1996

Aria: Ambienti di lavoro, Fiala in carbone attivo (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Chetoni: acetone, cicloesanone, metil isobutilchetone, 2-exanone	NIOSH 1300 1994
Compensi organici volatili (COV): acetato di etile, acetato di n-butile, alcool isobutilico, alcool n-butilico, alcool terz-butilico, cicloesano, n-epitano, n-esano, tetraidrofurano	OSHA 07 2000

Aria: Ambienti di lavoro, Fiala in gel di silice e membrana (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Acido fluoridrico, Acido cloridrico, Acido nitrico, Acido solforico, Acido fosforico, Acido bromidrico	NIOSH 7903 1994

Aria: Ambienti di lavoro, Membrana (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Alluminio, Antimonio, Bario, Boro, Calcio, Cromo, Ferro, Magnesio, Manganese, Nichel, Piombo, Rame, Sodio, Stagno, Zinco	NIOSH 7300 2003
Piombo	NIOSH 7105 1994

Aria: Ambiente di lavoro; Fiala carbone attivo (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Idrocarburi aromatici: benzene, toluene, o,m,p-xilene, cumene, viniltoluene, stirene, etilbenzene	NIOSH 1501 2003

Aria: Ambiente di lavoro; fiala attivata (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Glicoli: Glicole etilenico, glicole dietilenico, glicole trietilenico	NIOSH 5523 1996

Aria: ambienti di lavoro; membrana filtrante (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Polveri: frazione respirabile	M.U. 2010: 2011

Aria: Aria Ambiente; membrana filtrante (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Benzo(a) pirene	UNI EN 15549: 2008((escluso il par. 11.1))
Cadmio, Arsenico, Piombo, Nichel (nella frazione PM10 del particolato in sospensione)	UNI EN 14902:2005
Frazione massica PM 2,5	UNI EN 14907:2005

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A Revisione: 25 Data: 16/07/2013 Scheda 8 di 20 PA163AR25.pdf
--	---

Particolato in sospensione PM 10	UNI EN 12341:2001
Aria: emissioni in atmosfera	
Denominazione della prova / Campi di prova	Metodo di prova
Contenuto di vapor d'acqua del gas umido, Umidità	UNI EN 14790:2006
Fluoruri, Acido Fluoridrico (HF)	ISO 15713:2006
Particolato fine < 2,5 micron (PM 2,5), Particolato fine < 10 micron (PM 10)	ISO 23210: 2009
Aria: emissioni in atmosfera, Fiala in carbone attivo (nel caso di esclusione del campionamento)	
Denominazione della prova / Campi di prova	Metodo di prova
Sostanze organiche volatili (SOV): triclorometano (cloroformio), 1,1,1 tricloroetano (metilcloroformio), tetracloruro di carbonio, tricloroetilene, tetracloroetilene, 1,1,2,2 tetracloroetano, 1,2-dicloropropano, o-diclorobenzene, p-diclorobenzene, acetone, benzene, 2-butanon (MEK), n-butilacetato, butanolo, cumene, cicloesano, etilacetato, etilbenzene, eptano, esano, metilisobutiletano (MIBK), metilcicloesano, stirene, toluene, o,m-viniltoluene, xilene, p-metilstirene, cicloesanone, tetraidrofuranico, alcool terbutilico, alcool isobutilico, 1-butano	UNI EN 13649: 2002
Aria: emissioni in atmosfera, Filtro +Soluzione per campionamento (nel caso di esclusione del campionamento)	
Denominazione della prova / Campi di prova	Metodo di prova
Arsenico, Cadmio, Cromo, Cobalto, Rame, Manganese, Nichel, Piombo, Antimonio, Tallio, Vanadio	UNI EN 14385:2004
Mercurio totale	UNI EN 13211:2003 + UNI EN 1483:2008, UNI EN 1483:2008
Aria: emissioni in atmosfera, Membrana (nel caso di esclusione del campionamento)	
Denominazione della prova / Campi di prova	Metodo di prova
Polveri totali	UNI EN 13284-1: 2003
Solfuro di idrogeno	M.U. 634:84
Aria: emissioni in atmosfera, Membrana + soluzione di lavaggio (nel caso di esclusione del campionamento)	
Denominazione della prova / Campi di prova	Metodo di prova
Alluminio, Cadmio, Cromo, Manganese, Nichel, Piombo, Rame, Stagno, Zinco (su polveri)	UNI EN 13284-1:2003 + M.U. 723:86 + UNI EN ISO 11885:2009
Aria: emissioni in atmosfera, Soluzione per campionamento (nel caso di esclusione del campionamento)	
Denominazione della prova / Campi di prova	Metodo di prova
Acido cloridrico	UNI EN 1911: 2010 metodo C
Acido cloridrico, Acido fluoridrico	DM 25/08/2000 GU n° 223 23/09/2000 SO n° 158 All. 2 pag 37
Aldeidi: formaldeide, acetaldeide, propionaldeide, butirraldeide, benzaldeide	EPA 0011 1996 + EPA 8315A 1996
Ammoniaca	M.U. 632:84
Anidride solforosa, Biossido di zolfo	UNI EN 14791:2006 Metodo A
Aria: emissioni in atmosfera; materiale adsorbente+filtro+condensa (nel caso di esclusione del campionamento)	
Denominazione della prova / Campi di prova	Metodo di prova

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A
	Revisione: 25 Data: 16/07/2013
	Scheda 9 di 20 PA163AR25.pdf

Dibenzodiossine/furani policlorurati (PCDD/PCDF):
 2,3,7,8-Tetraclorodibenzodiossina (TCDD),
 1,2,3,7,8-Pentaclorodibenzodiossina (PeCDD),
 1,2,3,4,7,8-Esaclorodibenzodiossina (HxCDD),
 1,2,3,6,7,8-Esaclorodibenzodiossina (HxCDD),
 1,2,3,7,8,9-Esaclorodibenzodiossina (HxCDD),
 1,2,3,4,6,7,8-Eptaclorodibenzodiossina (HpCDD), Octaclorodibenzodiossina (OCDD) Policlorodibenzofurani (PCDF): 2,3,7,8-Tetraclorodibenzofurano (TCDF), 1,2,3,7,8-Pentaclorodibenzofurano (PeCDF),
 2,3,4,7,8-Pentaclorodibenzofurano (PeCDF),
 1,2,3,4,7,8-Esaclorodibenzofurano (HxCDF),
 1,2,3,6,7,8-Esaclorodibenzofurano (HxCDF),
 1,2,3,7,8,9-Esaclorodibenzofurano (HxCDF),
 2,3,4,6,7,8-Esaclorodibenzofurano (HxCDF),
 1,2,3,4,6,7,8-Eptaclorodibenzofurano (HpCDF), 1,2,3,4,7,8,9-Eptaclorodibenzofurano (HpCDF), Octaclorodibenzofurano (OCDF) -
 Sommatoria come Tossicità equivalente I-TEQ

UNI EN 1948-1: 2006 + UNI EN 1948-2: 2006 + UNI EN 1948-3:2006 + NATO /CCMS Report n° 176 1988

Aria: emissioni, Filtro + puff + resina + condensa (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Idrocarburi policiclici aromatici (IPA): fluorantene, benzo(a) antracene, crisene, benzo(b) fluorantene+ benzo(j)fluorantene, benzo(a) pirene, dibenzo(a,h) antracene, benzo(g,h,i) perilene, indeno(1,2,3,cd) pirene, benzo (k) fluorantene + benzo (j) fluorantene, dibenzo(a,e)pirene,dibenzo(a,h)pirene, dibenzo(a,i)pirene, dibenzo(a,l)pirene, dibenzo(a,h)pirene	ISO 11338-1: 2003 + ISO 11338-2: 2003, ISO 11338-2: 2003

Aria: emissioni, Soluzione per campionamento (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Ossidi di azoto, Ossidi di zolfo	DM 25/08/2000 GU n° 223 23/09/2000 SO n° 158 All 1 pag. 30

Aria:ambienti di lavoro; Membrana (nel caso di esclusione del campionamento)

Denominazione della prova / Campi di prova	Metodo di prova
Polveri totali, Polveri frazione inalabile	M.U.1998:05

Cacao e cioccolato

Denominazione della prova / Campi di prova	Metodo di prova
Umidità	DM 06/01/1979 GUSO n° 67 08/03/1979 All B Met 1 - A

Campioni ambientali incluse acque potabili, industriali, naturali e materiali associati come sedimenti, depositi e melme

Denominazione della prova / Campi di prova	Metodo di prova
Conta Legionella	ISO 11731: 1998

Carcasse comprese quelle di pollo

Denominazione della prova / Campi di prova	Metodo di prova
Carica microbica totale	ISO 17604:2003 /Amd 1: 2009 (escluso Annex D)+ ISO 4833:2003

Conta Enterobacteriaceae

Denominazione della prova / Campi di prova	Metodo di prova
	ISO 17604:2003 /Amd 1: 2009 (escluso Annex D) + ISO 21528-2:2004

Carne e derivati

Denominazione della prova / Campi di prova	Metodo di prova
Ceneri	AOAC 920.153 1920

Conta Pseudomonas spp presunto

Denominazione della prova / Campi di prova	Metodo di prova
	UNI EN ISO 13720: 2010

Umidità

Denominazione della prova / Campi di prova	Metodo di prova
	AOAC 950.46 Met B (a) 1991

Carni

Denominazione della prova / Campi di prova	Metodo di prova

LASER LAB Srl Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A
	Revisione: 25 Data: 16/07/2013
	Scheda 10 di 20 PA163AR25.pdf

Ricerca Trichinella	Reg CE 2075/2005 05/12/2005 GU CE L338 22/12/2005 Ali I Cap I (escluso par 2) + Reg CE 1245/2007 24/10/2007 GU CE L281 25/10/2007
Cereali e derivati	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Glutine secco	DM 23/07/1994 GU n° 186 10/08/1994 SO Pag 8
Ocratossina A (> 0,5 µg/kg)	MP 280/C rev 5 2013
pH	AOAC 943.02 1943
Sostanze azotate, Proteine (da calcolo)	DM 23/07/1994 GU SO n° 186 10/08/1994 Pag 2
Cereali e derivati (solo per sfarinati e pasta)	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Umidità	DM 27/05/1985 SO n° 3 GU n° 145 21/06/1985
Cereali e derivati, frutta, conserve vegetali, prodotti dell'industria dolciaria	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Fibra alimentare totale	AOAC 985.29 1986
Combustibili solidi non minerali ricavati da rifiuti (CDR), Non mineral refuse derived fuels (RDF)	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Piombo volatile	UNI EN 15402: 2011 + UNI EN 15411: 2011 Met. C + UNI EN ISO 11885:2009
Vetro	UNI 9903-14: 1997
Combustibili solidi secondari (CSS), Solid recovered fuels (SRF)	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Antimonio, Arsenico, Bario, Berillio, Cadmio, Cobalto, Cromo, Molibdeno, Manganese, Nichel, Piombo, Rame, Selenio, Tallio, Vanadio, Zinco, Mercurio	UNI EN 15411: 2011 Met. C + UNI EN ISO 11885:2009
Ceneri	UNI EN 15403: 2011
Umidità	UNI EN 15414-3: 2011
Concimi, Fertilizzanti, Compost, Ammendanti	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
pH	DM 19/07/1989 GU n° 196 23/07/1989 met. 4
Conserve vegetali	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Acidità totale, Acidità (come acido acetico) (da calcolo)	DM 03/02/1989 GU n° 168 20/07/1989 Met 15
pH	DM 03/02/1989 GU n° 168 20/07/1989 Met 17
Sodio cloruro	DM 03/02/1989 GU n° 168 20/07/1989 Met 33
Solidi totali/ Sostanza secca	DM 03/02/1989 GU n° 168 20/07/1989 Met 5
Umidità	DM 03/02/1989 GU n° 168 20/07/1989 Met 6
Fanghi , rifiuti	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Conducibilità in eluati da test di cessione in acqua	UNI 10802:2004 App A+ App B +UNI EN 12457-2:2004+ UNI EN 16192: 2012+ UNI EN 27888: 1995

LASER LAB SrL	Numero di accreditamento: 0142 Sede A
Via Custoza 31 66013 Chieti CH	Revisione: 25 Data: 16/07/2013
	Scheda 11 di 20 PA163AR25.pdf

Solventi organici alogenati: cloroformio , 1,2-dicloroetano, percloroetilene, tricloroetilene, 1,2-dicloropropano, metilcloroformio, 1,1,2,2-tetracloroetano, 1,1,2-tricloroetano, 1,2,3-tricloropropano, tetracloruro di carbonio.

EPA 3580A 1992 + EPA 8021B 1996

Solventi organici aromatici: benzene, toluene, o-xilene, m+p-xilene, etilbenzene, stirene, cumene , MTBE, m-Viniltoluene, p-vinilstirene, Acetone, 2-Butanone(MEK), 4-metil-2-pantanone (MIBK)

EPA 3580A 1992 + EPA 8015C 2007

Fanghi, Rifiuti

Denominazione della prova / Campi di prova

Dibenzodiossine/furani policlorurati (PCDD/PCDF):
2,3,7,8-Tetraclorodibenzodiossina (TCDD),
1,2,3,7,8-Pentaclorodibenzodiossina (PeCDD),
1,2,3,4,7,8-Esaclorodibenzodiossina (HxCDD),
1,2,3,6,7,8-Esaclorodibenzodiossina (HxCDD),
1,2,3,7,8,9-Esaclorodibenzodiossina (HxCDD),
1,2,3,4,6,7,8-Eptaclorodibenzodiossina (HpCDD), Octaclorodibenzodiossina (OCDD) Policlorodibenzofurani (PCDF): 2,3,7,8-Tetraclorodibenzofurano (TCDF), 1,2,3,7,8-Pentaclorodibenzofurano (PeCDF),
2,3,4,7,8-Pentaclorodibenzofurano (PeCDF),
1,2,3,4,7,8-Esaclorodibenzofurano (HxCDF),
1,2,3,6,7,8-Esaclorodibenzofurano (HxCDF),
1,2,3,7,8,9-Esaclorodibenzofurano (HxCDF),
2,3,4,6,7,8-Esaclorodibenzofurano (HxCDF),
1,2,3,4,6,7,8-Eptaclorodibenzofurano (HpCDF), 1,2,3,4,7,8,9-Eptaclorodibenzofurano (HpCDF), Octaclorodibenzofurano (OCDF) ;
Sommatoria PCDD/PCDF I-TEQ

Metodo di prova

EPA 1613 B 1994 +UNEP/POPS/COP.3/INF/27
11/04/2007

Fanghi, Rifiuti, Sedimenti, Terreni

Denominazione della prova / Campi di prova

Alluminio, Antimonio, Argento, Arsenico, Bario, Berillio, Boro, Cadmio, Calcio, Cobalto, Cromo, Ferro, Fosforo, Magnesio, Mercurio, Manganese, Molibdeno, Nichel, Piombo, Potassio, Rame, Selenio, Sodio, Stronzio, Tallio, Tellurio, Vanadio, Zinco, Stagno

Metodo di prova

EPA 3052 1996 + EPA 6010C 2007

Alluminio, Antimonio, Argento, Arsenico, Bario, Berillio, Boro, Cadmio, Calcio, Cobalto, Cromo, Ferro, Fosforo, Magnesio, Mercurio, Manganese, Molibdeno, Nichel, Piombo, Potassio, Rame, Selenio, Sodio, Stagno, Stronzio, Tallio, Tellurio, Vanadio, Zinco

EPA 3051A 2007 + EPA 6010C 2007

Alluminio, Antimonio, Argento, Bario, Berillio, Boro, Cadmio, Cobalto, Cromo, Ferro, Fosforo, Magnesio, Manganese, Mercurio, Molibdeno, Nichel, Piombo, Rame, Silicio, Stagno, Stronzio, Tallio, Tellurio, Vanadio, Zinco

EPA 3050B 1996 + EPA 6010C 2007

Amianto: polveri e fibre libere

CNR IRSA App III Q 64 Vol 3 1996

Azoto ammoniacale

CNR IRSA 7 Q 64 Vol 3 1986

Carbonio Organico Disciolto (DOC) in eluati da Test di cessione in acqua (a pH corretto 7,5-8,0)

UNI 10802:2004 App A+ App B + UNI EN 16192: 2012+ UNI CEN/TS 14429: 2006 + ISO 10523:2009 + UNI EN 1484:1999

Carbonio Organico Disciolto (DOC) in eluati da test di cessione in acqua

UNI 10802:2004 App A + App B + UNI EN 12457-2:2004+UNI EN 1484:1999

Carbonio Organico Totale (TOC)

UNI EN 13137:2002

Cianuri in eluati da test di cessione in acqua

UNI 10802:2004 App:A+App.B + UNI EN 12457-2:2004+UNI EN 16192:2012+M.U. 2251:2008

Cianuri liberi e totali

M.U. 2251:2008

Cloruri (Come Cl -), Solfati (Come SO4 2-), Fluoruri (Come F-), Nitrati (Come NO3-), Fosfati (Come PO4 3-)

EPA 9056A 2007

Cloruri, Solfati,Fluoruri in eluati da test di cessione in acqua

UNI 10802:2004 App A + App B + UNI EN 12457-2:2004 + UNI EN 16192: 2012 + UNI EN ISO 10304-1:2009

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A	
	Revisione: 25	Data: 16/07/2013
	Scheda 12 di 20	PA163AR25.pdf

Cromo esavalente (Cromo VI)	CNR IRSA 16 Q 64 Vol 3 1986
Densità	CNR IRSA 3 Q 64 Vol 2 1984
Elementi in eluati da test di cessione in acqua: Antimonio, Arsenico, Bario, Cadmio, Cromo, Molibdeno, Nichel, Piombo, Rame, Selenio, Zinco	UNI 10802:2004 App A + App B + UNI EN 12457-2:2004 + UNI EN ISO 11885:2009
Esteri dell'acido ftalico: Bis(2-etilesil)ftalato, butil benzil ftalato, di-n-butil ftalato, di-n-ottile ftalato, dietil ftalato, dimetil ftalato	EPA 3545A 2007 + EPA 8270D 2007
Idrocarburi C<12 (5<C<=12), Metil-tert-butiltere	EPA 3815 2007 + EPA 5021A 2003 + EPA 8015C 2007
Indice fenolo in eluati da test di cessione in acqua	UNI 10802:2004 App A + App B + UNI EN 16192: 2012 + ISO 6439:1990 met A
Mercurio	CNR IRSA 10 Q 64 Vol 3 1985 + APAT CNR IRSA 3200 A1 Man 29 2003
Mercurio in eluati da test di cessione in acqua	UNI 10802:2004 App A + App B + UNI EN 12457-2:2004 + UNI EN 1483:2008
Oli e grassi totali	CNR IRSA 21 Q 64 Vol 3 1988
Pesticidi organofosforati: Clorpirifos-me, Clorpirifos-et, Pirimifos-me , Paration-me, disulfoton,sulfotep	EPA 3550C 2007 + EPA 3640A 1994 + EPA 8270D 2007
Pesticidi organoclorurati : esaclorobenzene, a-HCH, b-HCH, aldrin, 4,4'-Diclorodifeniltricloroetano (4,4'-DDT), 2,4'-Diclorodifeniltricloroetano (2,4'-DDT), 4,4'-Diclorodifenildicloroetano (4,4'-DDD), 2,4'-Diclorodifenildicloroetano (2,4'-DDD), 4,4'-Diclorodifenildicloroetilene (4,4'-DDE), 2,4'-Diclorodifenildicloroetilene (2,4'-DDE), somma di DDD, DDT e DDE, lindano (g-HCH), dieldrin, isodrin, Endrin, Clordano.	
pH	CNR IRSA 1 Q 64 Vol 3 1985
pH in eluati da Test di cessione in acqua	UNI 10802:2004 App A+ App B + UNI EN 16192: 2012 + ISO 10523:2008
Piombo tetraetile	EPA 3545A 2007+ EPA 8270D 2007
Policlorobifenili (PCB): Aroclor 1260, Aroclor 1254, Aroclor 1242	CNR IRSA 24b Q 64 Vol 3 1988
Potere Calorifico	CNR IRSA 4 Q 64 Vol.2 1988
Solidi sospesi	CNR IRSA 1 Q 64 Vol.2 1984 A e B
Solidi Totali Disciolti (TDS) in eluati da Test di cessione in acqua	UNI 10802:2004 App A+ App B + UNI EN 12457-2:2004 + UNI EN 15216:2008
Solidi totali, Residuo secco a 105°C, Residuo secco a 550 °C	CNR IRSA 2 Q 64 Vol 2 1984
Fanghi, Rifiuti, Sedimenti, Terreni, Oli minerali, Combustibili solidi non minerali ricavati da rifiuti (CDR), Non mineral refuse derived fuels (RDF)	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Cloro post-combustione, Zolfo post-combustione	EPA 5050 1994 + EPA 9056A 2007
Fanghi, Rifiuti, Sedimenti, Terreni, substrati assorbenti da campionamenti di aria	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>

LASER LAB Srl Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A	
	Revisione: 25	Data: 16/07/2013
	Scheda 13 di 20	PA163AR25.pdf

Ammine aromatiche: anilina, o-anisidina, m-anisidina, p-anisidina, difenilammmina, p-toluidina

EPA 3545A 2007 + EPA 8270D 2007

Idrocarburi policiclici aromatici (IPA):

naftalene, acenaftilene, acenaftene, fluorene, fenantrene, antracene, fluorantene, pirene, dibenzo(a,e)pirene, dibenzo(a,l)pirene, dibenzo(a,i)pirene, dibenzo(a,h)pirene, benzo(a) antracene, crisene, benzo(b) fluorantene, benzo(a) pirene, dibenzo(a,h) antracene, benzo(g,h,i) perilene, indeno(1,2,3,cd) pirene, benzo(k) fluorantene; Clorobenzeni: 1,2,4-triclorobenzene, 1,2,4,5-tetraclorobenzene, pentaclorobenzene, esaclorobenzene;

Nitrobenzeni: Nitrobenzene, 2-cloronitrobenzene, 3-cloronitrobenzene, 4-cloronitrobenzene, 1,3-dinitrobenzene, 1,2-dinitrobenzene Fenoli: Fenolo, m+p cresolo, o-cresolo, 2-clorofenolo, 2,4-diclorofenolo, 2,4,6-triclorofenolo, Pentaclorofenolo.

Policlorobifenili (PCB): #28, #52, #95, #99, #101, #110, #128 + #138, #146, #149, #151, #153, #170, #177, #180, #183, #187; Policlorobifenili (PCB)

Diossina simile: #77, #81, #105, #114, #118, #123, #126, #156, #157, #167, #169, #189; PCB Totali.

Farina di grano tenero, semola di grano duro

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Filth Test	ISO 11050: 1993

Farine

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Ceneri	AOAC 923

Fertilizzanti, Compost, Ammendanti

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Umidità	DM 24/03/1986 Met B GU n° 180 05/08/1986

Frutta e vegetali

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
pH	ISO 1842:1991

Gas naturali e gas combustibili

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Caratteristiche fisiche calcolate a 15°C (288,15 K) e 1,01325 bar (101,325 Kpa) : Indice di Wobbe, densità, densità relativa	UNI EN 15984:2011 +UNI EN ISO 6976:2008 Par 8
Caratteristiche fisiche calcolate a 15°C (288,15 K) e 1,01325 bar (101,325 Kpa): Potere calorifico superiore,potere calorifico inferiore,peso molecolare medio	UNI EN 15984:2011 +UNI EN ISO 6976:2008 Par 5,6,7
Composizione centesimale: metano, etano, propano, butano, iso-butano, pentano, iso-pentano, esano, eptano, ottano, nonano, idrocarburi >C9 (decano), anidride carbonica, monossido di carbonio, ossigeno, azoto. Alcani come C, Tenore di carbonio	UNI EN 15984:2011

Dew point idrocarburi (da calcolo)	UNI EN 15984:2011+ ISO 23874:2006
------------------------------------	-----------------------------------

Fattore di compressione	UNI EN 15984:2011 +UNI EN ISO 6976:2008 Par 4.2
-------------------------	---

Fattore di emissione	UNI EN 15984:2011 DEC CE 156/2004 29/01/2004 GU CE L59 26/02/2004 All. 1 P.to 4.2.2.1.6
----------------------	---

Fattore di ossidazione	UNI EN 15984:2011 DEC CE 156/2004 29/01/2004 GU CE L59 26/02/2004 All. 2 P.to 2.1.1.1C
------------------------	--

Gas naturali e gas di discarica

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Composizione (%mol): Metano, etano, propano, butano, altri idrocarburi superiori (esano, pentano, eptano, ottano, idrocarburi >C9, (decano), anidride carbonica, elio, ossigeno, azoto, monossido di Carbonio, Alcani come C (tenore di Carbonio)	UNI EN ISO 10715:2001+ISO 6975:1997

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A	
	Revisione: 25	Data: 16/07/2013
	Scheda 14 di 20	PA163AR25.pdf

Solfuro di idrogeno(Acido solfidrico), tetraidrotiofene, dietilsolfuro, metilmercaptano, etilmercaptano, terbutilmercaptano, ossisolfuro di carbonio, zolfo da mercaptani (calcolo), zolfo totale (calcolo)

UNI EN ISO 19739: 2007/EC1: 2010

Latte

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Attività perossidasica	DM 26/03/1992 GUSO 90/1992 16/04/1992 Met. 3
Azoto non proteico	AOAC 991.21 1994
Azoto totale, Proteine (da calcolo)	AOAC 991.20 1994

Materiali da costruzione contenenti amianto

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Amianto	VDI 3866 Parte 1 Cap 6 : 2000+VDI 3866 Parte 2: 2001

Oli di oliva e oli di sansa

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Acidi grassi liberi (Acidità)	Reg CEE 2568/1991 11/07/1991 GU CEE L248 05/09/1991 All II Reg CE 702/2007 21/06/2007 GU CE L161 22/06/2007

Oli di oliva, oli di sansa, sostanze grasse

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Analisi spettrofotometrica nell'ultravioletto (escluso il passaggio su allumina) Delta K, K232, K270	Reg CEE 2568/1991 11/07/1991 GU CEE L248 05/09/1991 All IX Reg CEE 183/1993 29/01/1993 GU CEE L22 30/01/1993

Oli e grassi

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Saggio Kreiss	NGD C56 - 1979

Oli e grassi animali e vegetali

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Numero di Iodio	Reg CEE 2568/1991 11/07/1991 GU CEE L248 05/09/1991 All XVI
Numero di perossidi	Reg CEE 2568/1991 11/07/1991 GU CEE L248 05/09/1991 All III

Oli, Grassi, Cere

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Antimonio, Arsenico, Bario, Berillio, Cadmio, Cromo totale, Ferro, Manganese, Nichel, Piombo, Rame, Vanadio	EPA 3040A 1996 + EPA 6010C 2007

Olio d'oliva, oli vegetali per uso alimentare

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Polifenoli totali (> 50 mg/Kg come Acido Caffeico)	MP 67/C rev 10 2013

Olio di oliva e di sansa, Alimenti di origine vegetale, olii di oliva, di semi, e grassi vegetali

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Esteri metilici degli acidi grassi (composizione acidica)(Acido miristico (C 14:0), Acido palmitico (C 16:0), Acido palmitoleico (C 16:1), Acido eptadecanoico (C 17:0), Acido eptadecenoico (C 17:1), Acido stearico (C 18:0), Acido oleico (C 18:1), Acido linoleico (C 18:2), Acido arachido (C 20:0), Acido eicosenoico (C 20:1), Acido beenico (C 22:0), Acido lignocerico (C 24:0)).	Reg CE 796/2002 06/05/2002 GU CE L128 15/05/2002 All XB + Reg CEE 2568/1991 11/07/1991 GU CEE L248 05/09/1991 All XA+ Reg CEE 1429/1992 26/05/1992 GU CEE L150 02/06/1992

Pesce e derivati

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
---	------------------------

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A	
	Revisione: 25	Data: 16/07/2013
	Scheda 15 di 20	PA163AR25.pdf

Azoto Basico Volatile Totale (ABVT)	Reg CE 2074 2005 05/12/2005 GU CE L338 22/12/2005 All.II Cap. III
Pesce e frutti di mare	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Mercurio	UNI EN 13806:2003
Prodotti lattiero caseari	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Azoto solubile, proteine (da calcolo)	DM 21/04/1986 GU SO n° 229 02/10/1986 Met 8
Materia secca, umidità (da calcolo)	DM 21/04/1986 GU SO n° 229 02/10/1986 Met 2
pH	DM 21/04/1986 GU SO 229/1986 02/10/1986 Met 20
Sostanze azotate totali, proteine (da calcolo)	DM 21/04/1986 GU SO n° 229 02/10/1986 Met 6
Prodotti ortofrutticoli Alimenti non grassi: Drupacee (pesche, comprese nettarine e percoche, albicocche, ciliegie)	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Multiresiduale fitofarmaci: Metsulfuron-metile, Clorfenvinfos, Carbendazim, Fention, Malation, Pirimifos metile, Pirazofos, Acetamiprid, Benalaxil, Imidacloprid, Metalaxil, Pirimicarb, Tiametoxam, Tolclofos-metile, Nuarimolo, Fenarimolo, Fenitrotion, Diazinone, Metomil, Fenexamide, Esaconazolo, Imazalil, Mepanipirim, Clofentezina, Propizamide (Pronamide), Ometoato, Triadimefon, Triadimenol, Pirimetanil, Trifloxistrobina, Clorpirifos, Sistano (Miclobutanil), Penconazolo, Acefate, Tiram	UNI EN 12393-1:2009 + UNI EN 12393-2:2009 + UNI EN 12393-3:2009 met N
Prodotti ortofrutticoli: Ortaggi a foglia (lattughe e altre insalate, dolcetta, scarola, indivia, crescione, barbarea, rucola, foglie e germogli di cavolo)	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Ditiocarbammati (come CS2) (>0,25mg/kg)	MP 282/C rev 03 2013
Prodotti petroliferi ed olii usati e materiali correlati	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Policlorobifenili totali (PCB)	UNI EN 12766-1:2001 + UNI EN 12766-2:2004
Rifiuti	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Alluminio, Antimonio, Argento, Arsenico, Bario, Berillio, Boro, Cadmio, Calcio, Cobalto, Cromo, Ferro, Fosforo, Magnesio, Mercurio, Manganese, Molibdeno, Nichel, Piombo, Potassio, Rame, Selenio, Sodio, Stronzio, Tallio, Tellurio, Vanadio, Zinco, Zolfo, Stagno (espressi in diverse forme i.e. come ossidi...)	UNI EN 13657: 2004 + UNI EN ISO 11885: 2009
Arsenico, Bario, Berillio, Cadmio, Cobalto, Cromo, Nichel, Piombo, Rame, Selenio, Vanadio, Zinco su eluati da test cessione in acqua	DM 186/2006 05/04/2006 GU n°115 19/05/2006 All.3+ UNI 10802:2004 App.A +App.B+ UNI EN 16192: 2012 + UNI EN 12457-2:2004 + UNI EN ISO 11885:2009
Azoto nitrico, Cloruri, Solfati su eluati da test cessione in acqua	DM 186/2006 05/04/2006 GU n° 115 19/05/2006 All.3 + UNI 10802:2004 App.A + App.B+UNI EN 16192: 2012 + UNI EN 12457-2:2004 + UNI EN ISO 10304-1:2009
Ceneri (550°C)	UNI EN 15169: 2007
Cianuri su eluati da test cessione in acqua	DM 186/2006 05/04/2006 GU n°115 19/05/2006 All.3 + UNI 10802:2004 App.A+ App.B+ UNI EN 16192: 2012 + UNI EN 13370:2004 + M.U.2251:2008
Fluoruri su eluati da test di cessione in acqua	DM 186/2006 05/04/2006 GU n° 115 19/05/2006 All.3 + UNI 10802:2004 App A + App.B + UNI EN 12457-2:2004+ UNI EN 16192: 2012+ UNI EN ISO 10304-1:2009

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A
	Revisione: 25 Data: 16/07/2013
	Scheda 16 di 20 PA163AR25.pdf

Idrocarburi C10-C40 (Olio minerale C10-C40)	UNI EN 14039: 2005
Idrocarburi policiclici aromatici (IPA): naftalene, acenaftilene, acenaftene, fluorene, fenantrene, antracene, fluorantene, pirene, dibenzo(a,e)pirene, dibenzo(a,l)pirene, dibenzo(a,i)pirene, dibenzo(a,h)pirene, benzo(a) antracene, crisene, benzo(b) fluorantene, benzo(a) pirene, dibenzo(a,h) antracene, benzo(g,h,i)perilene, indeno(1,2,3,cd) pirene, benzo(k) fluorantene	UNI EN 15527: 2008
Mercurio su eluati da test cessione in acqua	DM 186/2006 05/04/2006 GU n°115 19/05/2006 All.3+ UNI 10802:2004 App.A + App: B+ UNI EN 12457-2:2004 + UNI EN 1483:2008
pH su eluati da test cessione in acqua	DM 186/2006 05/04/2006 GU 115 19/05/2006 All.3 + UNI 10802 2004 App.A + App.B+UNI EN 16192: 2012 + UNI EN 12457-2 2004 + ISO 10523:2008
Richiesta chimica di ossigeno (COD) su eluati da test cessione in acqua	DM 186/2006 05/04/2006 GU n°115 19/05/2006 All.3 + UNI 10802:2004 App.A + App.B+ UNI EN 16192: 2012 + UNI EN 12457-2:2004 + ISO 15705:2002
Sostanza secca (residuo a 105°C)	UNI EN 14346:2007 Met A
Rifiuti	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Arsenico, Berillio, Cadmio, Nichel, Piombo, Selenio, Vanadio su eluati da test cessione in acqua	DM 186/2006 05/04/2006 GU n°115 19/05/2006 All.3+ UNI 10802:2004 App.A + App.B+ UNI EN 12506:2004 + UNI EN 12457-2:2004 + UNI EN ISO 17294-2: 2005
Rifiuti, Sedimenti, Terreni	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Antimonio, Arsenico, Cadmio, Nichel, Piombo, Selenio su eluati da test di cessione in acqua	UNI 10802:2004 App A + App. B + UNI EN 12457-2:2004 + UNI EN ISO 17294-2: 2005
Suoli , Sedimenti, Terreni	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Atrazina, Alaclor (> 0,001 mg/Kg)	MP 277/C rev 5 2013
Suoli, Sedimenti, Terreni	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Acidità	DM 13/09/1999 SO n° 185 GU n° 248 21/10/1999 Met XIII.3 DM 25/03/2002 GU n° 84 10/04/2002
Azoto totale	DM 13/09/1999 SO n° 185 GU n° 248 21/10/1999 Met XIV.2 + XIV.3 DM 25/03/2002 GU n° 84 10/04/2002
Clorobenzeni: monoclorobenzene, 1,2-diclorobenzene, 1,4-diclorobenzene, 1,2,4-triclorobenzene, 1,2,4,5-tetraclorobenzene, pentaclorobenzene, esaclorobenzene.	EPA 3815 2007 + EPA 5021A 2003 + EPA 8021B 1996
Solventi organici clorurati e alogenati: cloroformio, tetrachloruro di carbonio, 1,1,1tricloroetano, tricloroetilene, tetrachloroetilene, 1,1,2,2tetrachloroetano, 1,1,1,2tetrachloroetano, clorometano, diclorometano, 1,2dicloroetano, esaclorobutadiene, 1,1dicloroetilene, 1,1dicloroetano, 1,2dicloropropano, 1,1,2tricloroetano, 1,2,3tricloropropano, 1,2dicloroetilene, 1,2 dibromoetano, monoclorobenzene, 1,2diclorobenzene, 1,4diclorobenzene, bromoformio, bromodicitrormetano, dibromoclorometano.	
Cloruri, Nitrati, Solfati, Floruri, Fosfati solubili in acqua	DM 13/09/1999 SO n° 185 GU n° 248 21/10/1999 Met IV.2 DM 25/03/2002 GU n° 84 10/04/2002
Conducibilità	DM 13/09/1999 SO n° 185 GU n° 248 21/10/1999 Met IV.1 + DM 25/03/2002 GU n° 84 10/04/2002

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A
	Revisione: 25 Data: 16/07/2013
Scheda 17 di 20	PA163AR25.pdf

Idrocarburi C>12, Idrocarburi C10-C40	ISO 16703: 2004
Indice di disponibilità per le piante dei metalli pesanti: Cadmio, Nichel, Piombo, Rame, Zinco	DM 13/09/1999 SO n° 185 GU n° 248 21/10/1999 Met XII.1, XII.2 + DM 25/03/2002 GU n° 84 10/04/2002
Metalli pesanti estraibili in acqua regia: Cadmio, Cobalto, Cromo, Manganese, Nichel, Piombo, Rame, Zinco	DM 13/09/1999 SO n° 185 GU n° 248 21/10/1999 Met XI.2 + DM 25/03/2002 GU n° 84 10/04/2002
Metalli pesanti estraibili in acqua regia: Cadmio, Cobalto, Cromo, Manganese, Nichel, Piombo, Rame, Zinco	DM 13/09/1999 SO n° 185 GU n° 248 21/10/1999 Met XI.1
pH	DM 13/09/1999 SO n° 185 GU n° 248 21/10/1999 Met III.1 + DM 25/03/2002 GU n° 84 10/04/2002
Scheletro (frazione granulometrica >= 2 mm; Terra fine (frazione granulometrica < 2 mm)	DM 13/09/1999 SO n° 185 GU n° 248 21/10/1999 Met II.1
Solidi disciolti nel suolo o negli estratti acquosi di suolo	DM 13/09/1999 SO n° 185 GU n° 248 21/10/1999 Met IV.3
Solventi organici aromatici: benzene, toluene, o-xilene, m+p-xilene, stilbenzene, stirene .	EPA 3815 2007 + EPA 5021A 2003 + EPA 8015C 2007
Solventi organici clorurati e alogenati: cloruro di vinile monomero (CVM), Cloroformio, Bromoformio, Dibromoclorometano, Bromodiclorometano	EPA 5035 1996 + EPA 8260C 2006
Umidità	DM 13/09/1999 GU n° 248 21/10/1999 SO n° 185 Met II.2
Superfici (Tamponi)	
<i>Denominazione della prova / Campi di prova</i>	
Conta di Enterobacteriaceae	<i>Metodo di prova</i> ISO 18593:2004 + ISO 21528-2:2004
Conta Coliformi	ISO 18593: 2004 + ISO 4832:2006
Conta Coliformi a 44°C	ISO 18593:2004 + NF V 08-060 2009
Conta Coliformi a 30°C	ISO 18593:2004 + NF V 08-050 2009
Conta di stafilococchi coagulasi positivi a 37 °C	ISO 18593 :2004 + NF V 08-057-1: 2004
Conta Escherichia coli beta-glucuronidasi positivo	ISO 18593: 2004 + ISO 16649-2:2001
Conta Lieviti e Muffe	ISO 18593 :2004 + NF V 08-059:2002
Conta Microrganismi a 30 °C	ISO 18593:2004 + ISO 4833:2003
Conta Pseudomonas spp presunto	ISO 18593 :2004 + UNI EN ISO 13720: 2010
Ricerca di Listeria monocytogenes	ISO18593:2004+ NF EN ISO 11290-1:1997/A1: 2005
Ricerca Listeria monocytogenes	ISO 18593:2004 + UNI EN ISO 11290-1:2005
Yogurt	
<i>Denominazione della prova / Campi di prova</i>	
Conta di microrganismi caratteristici	<i>Metodo di prova</i> Rapporti ISTISAN 1996/35 Met 24

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A
	Revisione: 25 Data: 16/07/2013
	Scheda 18 di 20 PA163AR25.pdf

ELENCO PROVE ACCREDITATE - CATEGORIA: II

Aria ambiente

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Biossido di zolfo (SO ₂)	UNI EN 14212:2012
Monossido di azoto (NO), Ossidi di Azoto (NO _x), Biossido di azoto (NO ₂),	UNI EN 14211:2012
Monossido di carbonio (CO)	UNI EN 14626:2012

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A
	Revisione: 25 Data: 16/07/2013
	Scheda 19 di 20 PA163AR25.pdf

ELENCO PROVE ACCREDITATE - CATEGORIA: III

Acque destinate al consumo umano e di piscina

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Cloro libero e Cloro totale	Rapporti ISTISAN 2007/31 pag 45 Met ISS BHD 038

Acque naturali

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Ossigeno disciolto	UNI EN ISO 5814: 2013

Acque naturali , di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Campionamento	APAT CNR IRSA 1030 Man 29 2003

Acque naturali (sotterranee, superficiali, di mare), acque di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Conducibilità elettrica	APAT CNR IRSA 2030 Man 29 2003

Acque naturali e di scarico, incluse acque di mare

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Temperatura	APAT CNR IRSA 2100 Man 29 2003

Acque naturali, sotterranee, superficiali, di mare, acque di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
pH	APAT CNR IRSA 2060 Man 29 2003
Potenziale Redox	APHA Standard Methods for the Examination of Water and Wastewater ed 22nd 2012 2580 B

Acque sotterranee

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Campionamento	M.U. 196/2:04

Acque superficiali, di scarico

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Campionamento	APAT CNR IRSA 6010 Man 29 2003

Ambiente abitativo ed esterno

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Livello continuo equivalente di pressione sonora ponderato A	DPCM 01/03/1991 GU n° 57 08/03/1991, L n° 447 26/10/1995 GU n° 254 30/10/1995 SO, DM 16/03/98 GU n° 76 01/04/98

Ambienti di lavoro

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Livello continuo equivalente di pressione sonora ponderato A (LAeq, T) Livello di pressione sonora di picco ponderato C Lpicco,C (ppeak)	ISO 1999:1990 + UNI EN ISO 9612:2011 +UNI 9432:2011
Livello di esposizione giornaliera al rumore (LEX, 8h) Livello di esposizione settimanale al rumore (LEX, w)	

Misurazione e Valutazione dell'esposizione dell'uomo alle vibrazioni trasmesse al corpo intero aw, A (8)

UNI ISO 2631-1:2008 + UNI EN 14253:2008

Misurazione e valutazione dell'esposizione dell'uomo alle vibrazioni trasmesse al sistema mano - braccio ahv, A(8)

UNI EN ISO 5349-1:2004 + UNI EN ISO 5349-2:2004

Aria: Emissioni in atmosfera

<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Biossido di Zolfo (SO2)	UNI 10393: 1995 (escluso il punto 7.2.1)

LASER LAB SrL Via Custoza 31 66013 Chieti CH	Numero di accreditamento: 0142 Sede A
	Revisione: 25 Data: 16/07/2013
	Scheda 20 di 20 PA163AR25.pdf

Biossido di zolfo(SO2), Ammoniaca(NH3), Acido fluoridrico(HF), Acido cloridrico(HCl), Ossido di azoto(NO), Biossido di azoto(NO2), Monossido di carbonio(CO), Biossido di carbonio(CO2), Acqua(H2O)	ASTM D6348-12
Carbonio Organico Totale (COT)	UNI EN 13526:2002
Carbonio Organico Totale (da 0 mg/m3 a 20 mg/m3) (COT)	UNI EN 12619:2002
Metano (CH4)	UNI EN ISO 25140: 2010
Monossido di Carbonio (CO)	UNI EN 15058:2006
Ossidi di Azoto (Nox), ossidi di azoto (come NO2), Monossido di azoto (NO), Biossido di azoto (NO2)	UNI EN 14792:2006
Ossigeno (O2)	UNI EN 14789:2006
Ossigeno (O2), Monossido di Carbonio (CO), Biossido di carbonio (CO2)	ISO 12039:2001 (escluso il punto 7.4)
Velocità, Portata, Temperatura, Pressione	UNI EN ISO 16911-1:2013 + UNI EN 15259:2008
Emissioni da combustione di gas naturale	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Monossido di Carbonio, Ossidi di Azoto (Nox), ossidi di azoto (come NO2), Monossido di azoto (NO), Ossigeno	ASTM D6522-11
Gas naturali e gas Combustibili	
<i>Denominazione della prova / Campi di prova</i>	<i>Metodo di prova</i>
Dew point acqua	ISO 6327:1981

Legenda

MP: Metodo di prova interno del laboratorio
AOAC: Association of Official Analytical Chemists
APAT: Agenzia per la Protezione dell'Ambiente e per i servizi Tecnici
CNR IRSA: Consiglio Nazionale delle Ricerche - Istituto di Ricerca sulle Acque
ASTM: American Society for Testing Materials
EPA: Environmental Protection Agency (USA)
Dec, Reg CEE: Decisione, Regolamento della Comunità Economica Europea
DLgs, DM, DPCM, OM: Decreto Legislativo, Decreto Ministeriale, Decreto del Presidente del Consiglio dei Ministri, Ordinanza Ministeriale della Repubblica italiana
UNI: Ente Nazionale di Unificazione Italiano
GU: Gazzetta Ufficiale
M.U.: Metodo UNICHIM (Associazione per l'unificazione nel settore dell'industria chimica)
EN: Norma Europea
ISO: International Organization for Standardization
NF: Norma AFNOR (Association Française de Normalisation)
NGD: Norme Grassi e Derivati
NIOSH: National Institute of Occupational Safety and Health
OSHA: Occupation Safety and Health Administration
ISTISAN: Istituto Superiore di Sanità

ACCREDIA
Il Direttore del Dipartimento
(Dr. Paolo Bianco)

Firmato digitalmente da Blanco Paolo
ND: c=IT, o=ACREDIA/10566361001, cn=Blanco
Paolo, serialNumber=f3:NCPLA52M23L219N,
givenName=Paolo, sn=Blanco,
dnQualifier=11004771, title=Direttore
Dipartimento Laboratori di prova
Data: 2013.07.24 20:25:43 +02'00'

Bianco Paolo